

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2733

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136657

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136657BS							
BOD5	mg/L	198	192		97	1	84.6-115.4	07/30/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q2733

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB136662

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136662BS							
TSS	mg/L	550	533		97	1	90-110	07/31/2025