



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2733

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB55	Mercury	0.076	+/-0.2	U	0.20	CV	08/06/2025	14:10	LB136725

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2733

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	08/06/2025	14:14	LB136725
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	08/06/2025	15:06	LB136725
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	08/06/2025	15:26	LB136725

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2733

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	11:52	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	11:52	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	11:52	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	11:52	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	11:52	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q2733

Contract: ENTA05

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	12:29	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	12:29	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	12:29	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	12:29	LB136697
CCB02	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	13:41	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	13:41	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	13:41	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	13:41	LB136697
CCB03	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	14:54	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	14:54	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	14:54	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	14:54	LB136697
CCB04	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	15:53	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	15:53	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	15:53	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	15:53	LB136697
CCB05	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	16:55	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	16:55	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	16:55	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	16:55	LB136697
CCB06	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	17:45	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	17:45	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	17:45	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	17:45	LB136697
CCB07	Cadmium	0.50	+/-3	U	6.00	P	08/04/2025	18:02	LB136697
	Copper	4.60	+/-10	U	20.0	P	08/04/2025	18:02	LB136697
	Lead	2.30	+/-6	U	12.0	P	08/04/2025	18:02	LB136697
	Nickel	3.06	+/-20	U	40.0	P	08/04/2025	18:02	LB136697
	Zinc	16.7	+/-20	U	40.0	P	08/04/2025	18:02	LB136697