

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>ENTACT</u>	SDG No.:	<u>Q2733</u>
Contract:	<u>ENTA05</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.56	1.0	56	-5	7	08/04/2025	12:01	LB136697
	Copper	10.5	2.0	524	-18	22	08/04/2025	12:01	LB136697
	Lead	-8.46			-12	12	08/04/2025	12:01	LB136697
	Nickel	4.92	2.0	246	-38	42	08/04/2025	12:01	LB136697
	Zinc	0.33			-40	40	08/04/2025	12:01	LB136697
ICSAB01	Cadmium	983	972	101	826	1120	08/04/2025	12:05	LB136697
	Copper	490	511	96	434	588	08/04/2025	12:05	LB136697
	Lead	37.5	49.0	76	37	61	08/04/2025	12:05	LB136697
	Nickel	988	954	104	810	1100	08/04/2025	12:05	LB136697
	Zinc	1020	952	107	809	1095	08/04/2025	12:05	LB136697