

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>ENTACT</u>	level: <u>low</u>	sdg no.: <u>Q2733</u>
contract: <u>ENTA05</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2733-01</u>	client id: <u>TW-WTS-12MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2733-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	100		3.00	U	100	100		P
Copper	ug/L	75 - 125	153		10.0	U	150	102		P
Lead	ug/L	75 - 125	482		6.00	U	500	96		P
Nickel	ug/L	75 - 125	259		2.42	J	250	103		P
Zinc	ug/L	75 - 125	114		20.0	U	100	114		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>ENTACT</u>	level: <u>low</u>	sdg no.: <u>Q2733</u>
contract: <u>ENTA05</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2733-01</u>	client id: <u>TW-WTS-12MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2733-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cadmium	ug/L	75 - 125	101		3.00	U	100	101		P
Copper	ug/L	75 - 125	154		10.0	U	150	103		P
Lead	ug/L	75 - 125	484		6.00	U	500	97		P
Nickel	ug/L	75 - 125	261		2.42	J	250	103		P
Zinc	ug/L	75 - 125	114		20.0	U	100	114		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: ENTACT level: low sdg no.: Q2733
contract: ENTA05 lab code: ACE
matrix: Water sample id: Q2756-01 client id: 71725MS
Percent Solids for Sample: NA Spiked ID: Q2756-01MS Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.42		0.20	U	4.0	86		CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: ENTACT **level:** low **sdg no.:** Q2733
contract: ENTA05 **lab code:** ACE
matrix: Water **sample id:** Q2756-01 **client id:** 71725MSD
Percent Solids for Sample: NA **Spiked ID:** Q2756-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	3.21		0.20	U	4.0	80		CV