

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2733</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2733-01</u>	Client ID:	<u>TW-WTS-12DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2733-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	3.00	U	3.00	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Lead	ug/L	20	6.00	U	6.00	U			P
Nickel	ug/L	20	2.42	J	2.65	J	9		P
Zinc	ug/L	20	20.0	U	20.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: <u>ENTACT</u>	Level: <u>LOW</u>	SDG No.: <u>Q2733</u>
Contract: <u>ENTA05</u>		Lab Code: <u>ACE</u>
Matrix: <u>Water</u>	Sample ID: <u>Q2733-01MS</u>	Client ID: <u>TW-WTS-12MSD</u>
Percent Solids for Sample: <u>NA</u>	Duplicate ID <u>Q2733-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cadmium	ug/L	20	100		101		1		P
Copper	ug/L	20	153		154		1		P
Lead	ug/L	20	482		484		0		P
Nickel	ug/L	20	259		261		1		P
Zinc	ug/L	20	114		114		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2733</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2756-01</u>	Client ID:	<u>71725DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2756-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2733</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2756-01MS</u>	Client ID:	<u>71725MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2756-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.42		3.21		6		CV