

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ENTA05
SDG NO.: Q2733

Calibration Date(s): 07/23/2025 07/23/2025
Calibration Times: 11:32 19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO112414.D</u>		CF 750 = <u>PO112415.D</u>			
CF 500 = <u>PO112416.D</u>		CF 250 = <u>PO112417.D</u>		CF 050 = <u>PO112418.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	281481703	285148303	290849518	279501268	204848240	268365806
Aroclor-1016-2	(2)	418001550	431852928	433849136	420704900	326178800	406117463
Aroclor-1016-3	(3)	263638281	273794351	273610504	262887848	234438940	261673985
Aroclor-1016-4	(4)	216165184	219268720	223332708	213740540	186700200	211841470
Aroclor-1016-5	(5)	221113370	228268088	233867428	235369996	178734160	219470608
Aroclor-1260-1	(1)	431773857	450001749	476240286	447612572	391026880	439331069
Aroclor-1260-2	(2)	673380136	698254531	722944530	686506064	658383800	687893812
Aroclor-1260-3	(3)	611274773	627246767	610311884	614713140	476551180	588019549
Aroclor-1260-4	(4)	447285263	461047633	473248904	505121600	368319820	451004644
Aroclor-1260-5	(5)	1262689005	1283961461	1337121492	1347466800	964454900	1239138732
Decachlorobiphenyl		7456684020	7676252293	7829495380	7672255680	5924316000	7311800675
Tetrachloro-m-xylene		8554863660	8750848067	8754072420	8314620000	6283438800	8131568589
Aroclor-1242-1	(1)	256299409	259568919	260660454	265097176	196334180	247592028
Aroclor-1242-2	(2)	381137122	387088803	391556862	393536916	295648160	369793573
Aroclor-1242-3	(3)	241769617	246476011	248315054	252195680	194532760	236657824
Aroclor-1242-4	(4)	195717021	198363989	201564902	203398396	172348620	194278586
Aroclor-1242-5	(5)	215656431	220174447	229182990	229543836	175020200	213915581
Decachlorobiphenyl		8000645330	8037989413	8087967020	8086832920	6389983800	7720683697
Tetrachloro-m-xylene		9125652230	9195232120	9219491440	9207437080	6341808000	8617924174
Aroclor-1248-1	(1)	189314500	193406379	198364040	203445536	151513120	187208715
Aroclor-1248-2	(2)	257058197	266019899	270572928	278752472	203158020	255112303
Aroclor-1248-3	(3)	329824745	342687100	350417966	364320620	331255900	343701266
Aroclor-1248-4	(4)	503150480	523568551	540549606	563745892	409374460	508077798
Aroclor-1248-5	(5)	346508966	363261911	374337920	392493440	271353380	349591123
Decachlorobiphenyl		7900018490	8086308400	8195171040	8262080520	6143274400	7717370570
Tetrachloro-m-xylene		8946590960	9189061507	9273186240	9357342160	6567561400	8666748453
Aroclor-1254-1	(1)	525205609	551869540	568048078	599161796	500508140	548958633
Aroclor-1254-2	(2)	467389736	490253857	501431616	531267632	439250900	485918748
Aroclor-1254-3	(3)	736681841	765368536	778116616	788008044	653257120	744286431
Aroclor-1254-4	(4)	537295054	560529123	571244926	572391952	507057560	549703723
Aroclor-1254-5	(5)	713157717	729355651	738947588	743850736	613809420	707824222
Decachlorobiphenyl		7810644110	7988158653	8118440840	8069654680	6321650600	7661709777
Tetrachloro-m-xylene		8477893750	8928798520	9019842540	9077508800	7004909800	8501790682
Aroclor-1268-1	(1)	1805209287	1836925397	1850534848	1852730580	1410311920	1751142406

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	1523811042	1548194379	1559255960	1566806172	1167050180	1473023547	11
Aroclor-1268-3	(3)	1298638460	1316156257	1333295814	1329215988	983852920	1252231888	11
Aroclor-1268-4	(4)	502325303	515160189	514467028	515975792	361392800	481864222	13
Aroclor-1268-5	(5)	3537850645	3568750916	3561527908	3536594816	2581136540	3357172165	12
Decachlorobiphenyl		14538757830	14683927053	14777305700	14801547680	10962984400	13952904533	12
Tetrachloro-m-xylene		8916461140	9229914440	9271431040	9355148600	6516020600	8657795164	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance
Lab Code: ACE
Instrument ID: ECD_O

Contract: ENTA05
SDG NO.: Q2733

Calibration Date(s): 07/23/2025 07/23/2025
Calibration Times: 11:32 19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PO112414.D</u>		CF 750 = <u>PO112415.D</u>			
CF 500 = <u>PO112416.D</u>		CF 250 = <u>PO112417.D</u>		CF 050 = <u>PO112418.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	171399488	177928739	180256326	179838608	165612060	175007044 3
Aroclor-1016-2	(2)	254283756	266059885	273532534	272267680	236331280	260495027 6
Aroclor-1016-3	(3)	134555314	139369485	142347070	141007248	122553240	135966471 6
Aroclor-1016-4	(4)	106915956	111683577	115661150	118698688	102464540	111084782 6
Aroclor-1016-5	(5)	139169322	144750759	147937312	152336452	132321220	143303013 5
Aroclor-1260-1	(1)	240286237	254043527	261134446	258041412	229050000	248511124 5
Aroclor-1260-2	(2)	317223235	329626419	335041102	340620804	324956000	329493512 3
Aroclor-1260-3	(3)	256241072	262701627	269297872	275910760	220454080	256921082 8
Aroclor-1260-4	(4)	185740526	193469844	197255514	200936168	179151380	191310686 4
Aroclor-1260-5	(5)	419199446	421594747	433075366	434815624	352283180	412193673 8
Decachlorobiphenyl		1753295660	1820833173	1877801640	1892185680	1491444000	1767112031 9
Tetrachloro-m-xylene		5169666700	5314856493	5330554580	5111206560	3919594600	4969175787 12
Aroclor-1242-1	(1)	153645977	158282413	162201976	169125204	148714620	158394038 5
Aroclor-1242-2	(2)	230553136	236233123	241816698	253231308	206497880	233666429 7
Aroclor-1242-3	(3)	120997871	124593004	126633224	130347852	103373540	121189098 8
Aroclor-1242-4	(4)	116921071	120755713	124229780	129097232	122693120	122739383 3
Aroclor-1242-5	(5)	154482444	160907067	167116390	163960092	158390400	160971279 3
Decachlorobiphenyl		1831604040	1871840933	1924104720	1967119920	1579850400	1834904003 8
Tetrachloro-m-xylene		5483238770	5543459253	5573552360	5630865680	4061456200	5258514453 13
Aroclor-1248-1	(1)	116797575	121746903	126835386	132022196	110144520	121509316 7
Aroclor-1248-2	(2)	161755092	168905528	175846314	185366264	154525280	169279696 7
Aroclor-1248-3	(3)	170672394	178220732	184519824	194073912	160748960	177647164 7
Aroclor-1248-4	(4)	202462025	210645355	219142380	229572292	179950700	208354550 9
Aroclor-1248-5	(5)	207118457	216119221	227433754	242834452	180319280	214765033 10
Decachlorobiphenyl		1836524820	1878481893	1941164520	1982063240	1517022000	1831051295 10
Tetrachloro-m-xylene		5362990180	5502770333	5558036420	5629888520	4177350800	5246207251 12
Aroclor-1254-1	(1)	308086628	324598995	339171148	360746064	310502200	328621007 6
Aroclor-1254-2	(2)	269579969	283423153	295499466	318260532	280690420	289490708 6
Aroclor-1254-3	(3)	411643535	427646563	444183030	462345912	391487940	427461396 6
Aroclor-1254-4	(4)	256982627	269810957	279068642	286892132	272060980	272963068 4
Aroclor-1254-5	(5)	324913084	333876627	346327314	357344936	307636300	334019652 5
Decachlorobiphenyl		1794065640	1872151533	1910898600	1948005080	1573949600	1819814091 8
Tetrachloro-m-xylene		5135321880	5340323933	5419556800	5524764520	4404208000	5164835027 9
Aroclor-1268-1	(1)	514567345	516695037	532735372	551455976	473857920	517862330 5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Aroclor-1268-2	(2)	421574479	425971267	436618852	454509592	386645480	425063934	6
Aroclor-1268-3	(3)	318758782	324505707	331241876	345977680	285851460	321267101	7
Aroclor-1268-4	(4)	111516853	115705837	114796592	121616720	92841800	111295560	9
Aroclor-1268-5	(5)	742715083	756736297	758808540	780432416	619995020	731737471	8
Decachlorobiphenyl		3224484190	3300929360	3351395200	3481054400	2726789400	3216930510	9
Tetrachloro-m-xylene		5326223920	5511724707	5541913900	5672741280	4190979000	5248716561	12