

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2733
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112699.D **Time Analyzed:** 05:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.655	4.855	527.860	500.000	5.6
Aroclor-1016-2	4.772	4.674	4.874	527.500	500.000	5.5
Aroclor-1016-3	4.829	4.731	4.931	523.700	500.000	4.7
Aroclor-1016-4	4.949	4.851	5.051	523.450	500.000	4.7
Aroclor-1016-5	5.204	5.108	5.308	545.960	500.000	9.2
Aroclor-1260-1	6.242	6.145	6.345	508.960	500.000	1.8
Aroclor-1260-2	6.431	6.333	6.533	496.200	500.000	-0.8
Aroclor-1260-3	6.797	6.700	6.900	570.430	500.000	14.1
Aroclor-1260-4	7.056	6.959	7.159	562.600	500.000	12.5
Aroclor-1260-5	7.299	7.203	7.403	509.040	500.000	1.8
Decachlorobiphenyl	8.689	8.593	8.793	46.740	50.000	-6.5
Tetrachloro-m-xylene	3.669	3.569	3.769	51.450	50.000	2.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q2733
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/02/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112699.D **Time Analyzed:** 05:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.734	4.638	4.838	487.380	500.000	-2.5
Aroclor-1016-2	4.753	4.656	4.856	484.740	500.000	-3.1
Aroclor-1016-3	4.927	4.831	5.031	488.910	500.000	-2.2
Aroclor-1016-4	4.970	4.873	5.073	482.960	500.000	-3.4
Aroclor-1016-5	5.182	5.085	5.285	524.090	500.000	4.8
Aroclor-1260-1	6.210	6.113	6.313	508.980	500.000	1.8
Aroclor-1260-2	6.398	6.302	6.502	503.840	500.000	0.8
Aroclor-1260-3	6.548	6.453	6.653	506.700	500.000	1.3
Aroclor-1260-4	7.018	6.923	7.123	495.640	500.000	-0.9
Aroclor-1260-5	7.260	7.164	7.364	505.010	500.000	1.0
Decachlorobiphenyl	8.634	8.540	8.740	49.090	50.000	-1.8
Tetrachloro-m-xylene	3.660	3.562	3.762	49.390	50.000	-1.2