

Surrogate Summary

SW-846

SDG No.: Q2733

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169087BL	PB169087BL	2-Fluorophenol	150	110	73		15 (23)	110 (138)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	79.7	80		30 (67)	130 (132)
		2-Fluorobiphenyl	100	76.5	76		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	87		15 (44)	110 (137)
		Terphenyl-d14	100	84.3	84		30 (42)	130 (152)
PB169087BS	PB169087BS	2-Fluorophenol	150	120	80		15 (23)	110 (138)
		Phenol-d6	150	122	81		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.2	78		30 (67)	130 (132)
		2-Fluorobiphenyl	100	82.8	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	138	92		15 (44)	110 (137)
		Terphenyl-d14	100	88.6	89		30 (42)	130 (152)
PB169087BSD	PB169087BSD	2-Fluorophenol	150	131	87		15 (23)	110 (138)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	84.9	85		30 (67)	130 (132)
		2-Fluorobiphenyl	100	81.8	82		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	135	90		15 (44)	110 (137)
		Terphenyl-d14	100	100	100		30 (42)	130 (152)
Q2733-01	TW-WTS-12	2-Fluorophenol	150	57.3	38		15 (23)	110 (138)
		Phenol-d6	150	42.8	29		15 (10)	110 (134)
		Nitrobenzene-d5	100	82.2	82		30 (67)	130 (132)
		2-Fluorobiphenyl	100	72.8	73		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	109	72		15 (44)	110 (137)
		Terphenyl-d14	100	68.9	69		30 (42)	130 (152)