

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q2733
 Lab File ID: VX047179.D Date Analyzed: 07/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	284486	5.56	498230	6.76	431603	10.05
UPPER LIMIT	568972	6.056	996460	7.263	863206	10.549
LOWER LIMIT	142243	5.056	249115	6.263	215802	9.549
EPA SAMPLE NO.						
TW-WTS-12	351724	5.56	588680	6.76	536499	10.05
VX0730WBL01	349455	5.56	583262	6.76	517110	10.06
VX0730WBS01	298174	5.56	503453	6.76	446612	10.06
VX0730WBSD01	271218	5.56	470951	6.76	421865	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.