

DATA PACKAGE

GENERAL CHEMISTRY
METALS

PROJECT NAME : BANKER

G ENVIRONMENTAL

8 Carriage Ln

Succasunna, NJ - 07876

Phone No: 973-294-1771

ORDER ID : Q2742

ATTENTION : Gary Landis



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	5
2.1) Metals-AES- Case Narrative	5
2.2) Genchem- Case Narrative	6
3) Qualifier Page	7
4) QA Checklist	8
5) Metals-AES Data	9
6) Genchem Data	46
7) Shipping Document	60
7.1) CHAIN OF CUSTODY	61
7.2) Lab Certificate	62

1
2
3
4
5
6
7

DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

1

Laboratory Name : Alliance Technical Group LLC Client : G Environmental

Project Location : NJ Project Number : Banker

Laboratory Sample ID(s) : Q2742 Sampling Date(s) : 7/31/2025

List DKQP Methods Used (e.g., 8260,8270, et Cetra) **300.0,6010D,SOP**

1	For each analytical method referenced in this laboratory report package, were all specified QA/QC performance criteria followed, including the requirement to explain any criteria falling outside of acceptable guidelines, as specified in the NJDEP Data of Known Quality performance standards?	☒ Yes ☐ No
1A	Were the method specified handling, preservation, and holding time requirements met?	☒ Yes ☐ No
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)	☐ Yes ☐ No ☒ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?	☒ Yes ☐ No
3	Were samples received at an appropriate temperature (4±2° C)?	☒ Yes ☐ No ☐ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?	☐ Yes ☒ No
5	a)Were reporting limits specified or referenced on the chain-of-custody or communicated to the laboratory prior to sample receipt? b)Were these reporting limits met?	☒ Yes ☐ No ☒ Yes ☐ No ☐ N/A
6	For each analytical method referenced in this laboratory report package, were results reported for all constituents identified in the method-specific analyte lists presented in the DKQP documents and/or site-specific QAPP?	☒ Yes ☐ No
7	Are project-specific matrix spikes and/or laboratory duplicates included in this data set?	☐ Yes ☒ No

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."

Cover Page

Order ID : Q2742

Project ID : Banker

Client : G Environmental

Lab Sample Number

Q2742-01

Client Sample Number

9 Banker

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:06 pm, Aug 07, 2025

Date: 8/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: Banker

Project # N/A

Order ID # Q2742

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

1 Water sample was received on 07/31/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Anions Group1 and Metals Group4. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:06 pm, Aug 07, 2025



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

G Environmental

Project Name: Banker

Project # N/A

Order ID # Q2742

Test Name: Anions Group1

A. Number of Samples and Date of Receipt:

1 Water sample was received on 07/31/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Anions Group1. This data package contains results for Anions Group1.

C. Analytical Techniques:

The analysis of Anions Group1 was based on method 300.0.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample 9 Banker was diluted due to high concentrations for Chloride.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (9 BankerMS) analysis met criteria for all compounds except for Anions Group1(Chloride) due to matrix interference.

The Matrix Spike Duplicate (9 BankerMSD) analysis met criteria for all compounds except for Anions Group1(Chloride) due to matrix interference.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:07 pm, Aug 07, 2025

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2742

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 08/06/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q2742	Order ID:	Q2742
Client:	G Environmental	Project ID:	Banker

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : 9 Banker								
Q2742-01	9 Banker	Water	Sodium	271000		434	1000	ug/L



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	07/31/25
Project:	Banker	Date Received:	07/31/25
Client Sample ID:	9 Banker	SDG No.:	Q2742
Lab Sample ID:	Q2742-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	271000	1	434		1000	ug/L	08/01/25 10:05	08/04/25 14:12	6010D	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>G Environmental</u>	SDG No.:	<u>Q2742</u>
Contract:	<u>GENV01</u>	Lab Code:	<u>ACE</u>

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	868	+/-1000	U	2000	P	08/04/2025	11:52	LB136697

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental **SDG No.:** Q2742
Contract: GENV01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	868	+/-1000	U	2000	P	08/04/2025	12:29	LB136697
CCB02	Sodium	868	+/-1000	U	2000	P	08/04/2025	13:41	LB136697
CCB03	Sodium	868	+/-1000	U	2000	P	08/04/2025	14:54	LB136697
CCB04	Sodium	868	+/-1000	U	2000	P	08/04/2025	15:53	LB136697
CCB05	Sodium	868	+/-1000	U	2000	P	08/04/2025	16:55	LB136697
CCB06	Sodium	868	+/-1000	U	2000	P	08/04/2025	17:45	LB136697
CCB07	Sodium	868	+/-1000	U	2000	P	08/04/2025	18:02	LB136697

A
B
C
D
E
F
G
H
I
J

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2742

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169078BL		WATER		Batch Number:	PB169078		Prep Date:	08/01/2025	
	Sodium	434	<500	U	1000	P	08/04/2025	13:29	LB136697

A
B
C
D
E
F
G
H
I
J



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19500	20000	98	90 - 110	P	08/04/2025	11:41	LB136697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental
Contract: GENV01
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

SDG No.: Q2742
Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1990	2000	100	80 - 120	P	08/04/2025	11:45	LB136697

A

B

C

D

E

F

G

H

I

J

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	23900	25000	96	90 - 110	P	08/04/2025	12:25	LB136697
CCV02	Sodium	24500	25000	98	90 - 110	P	08/04/2025	13:37	LB136697
CCV03	Sodium	23600	25000	94	90 - 110	P	08/04/2025	14:43	LB136697
CCV04	Sodium	23200	25000	93	90 - 110	P	08/04/2025	15:48	LB136697
CCV05	Sodium	22900	25000	92	90 - 110	P	08/04/2025	16:51	LB136697
CCV06	Sodium	24600	25000	98	90 - 110	P	08/04/2025	17:41	LB136697
CCV07	Sodium	24600	25000	98	90 - 110	P	08/04/2025	17:58	LB136697



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Sodium	2030	2000	102	65 - 135	P	08/04/2025	11:57	LB136697

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: G Environmental **SDG No.:** Q2742
Contract: GENV01 **Lab Code:** ACE
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Sodium	115			0	0	08/04/2025	12:01	LB136697
ICSAB01	Sodium	94.2			0	0	08/04/2025	12:05	LB136697



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2742</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2733-01</u>	client id: <u>TW-WTS-12MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2733-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	75800		73800		1500	134		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q2742</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2733-01</u>	client id: <u>TW-WTS-12MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2733-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	71600		73800		1500	-147		P

A

B

C

D

E

F

G

H

I

J

Metals
- 5b -

Client:G Environmental

SDG No.:Q2742

Contract:GENV01

Lab Code:ACE

Matrix:

Level:LOW

Client ID:

Sample ID:

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

A

B

C

D

E

F

G

H

I

J

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2742</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2733-01</u>	Client ID:	<u>TW-WTS-12DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2733-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	73800		77200		5		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2742</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2733-01MS</u>	Client ID:	<u>TW-WTS-12MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2733-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	75800		71600		6		P

“A control limit of $\pm 20\%$ RPD for each matrix applies for sample values greater than 10 times Detection Limit”

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client:	<u>G Environmental</u>	SDG No.:	<u>Q2742</u>
Contract:	<u>GENV01</u>	Lab Code:	<u>ACE</u>

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169078BS Sodium	ug/L	1500	1430		95	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TW-WTS-12L

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **Lb No.:** lb136697 **Lab Sample ID :** Q2733-01L **SDG No.:** Q2742
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Sodium	73800	66600	10		P

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q2742

Instrument id number: _____

Method: _____

Run number: LB136697

Start date: 08/04/2025

End date: 08/04/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1115	Na
S1	S1	1	1120	Na
S2	S2	1	1124	Na
S3	S3	1	1128	Na
S4	S4	1	1132	Na
S5	S5	1	1136	Na
ICV01	ICV01	1	1141	Na
LLICV01	LLICV01	1	1145	Na
ICB01	ICB01	1	1152	Na
CRI01	CRI01	1	1157	Na
ICSA01	ICSA01	1	1201	Na
ICSAB01	ICSAB01	1	1205	Na
CCV01	CCV01	1	1225	Na
CCB01	CCB01	1	1229	Na
PB169078BL	PB169078BL	1	1329	Na
PB169078BS	PB169078BS	1	1333	Na
CCV02	CCV02	1	1337	Na
CCB02	CCB02	1	1341	Na
Q2733-01DUP	TW-WTS-12DUP	1	1350	Na
Q2733-01L	TW-WTS-12L	5	1354	Na
Q2733-01MS	TW-WTS-12MS	1	1359	Na
Q2733-01MSD	TW-WTS-12MSD	1	1403	Na
Q2742-01	9 Banker	1	1412	Na
CCV03	CCV03	1	1443	Na
CCB03	CCB03	1	1454	Na
CCV04	CCV04	1	1548	Na
CCB04	CCB04	1	1553	Na
CCV05	CCV05	1	1651	Na
CCB05	CCB05	1	1655	Na
CCV06	CCV06	1	1741	Na
CCB06	CCB06	1	1745	Na
CCV07	CCV07	1	1758	Na
CCB07	CCB07	1	1802	Na



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

LAB CHRONICLE

OrderID:	Q2742	OrderDate:	7/31/2025 3:56:00 PM
Client:	G Environmental	Project:	Banker
Contact:	Gary Landis	Location:	O22

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2742-01	9 Banker	Water	Metals Group4	6010D	07/31/25	08/01/25	08/04/25	07/31/25



METAL PREPARATION & ANALYICAL SUMMARY

Metals
- 13 -

SAMPLE PREPARATION SUMMARY

Client: G Environmental

SDG No.: Q2742

Contract: GENV01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169078							
PB169078BL	PB169078BL	MB	WATER	08/01/2025	50.0	25.0	
PB169078BS	PB169078BS	LCS	WATER	08/01/2025	50.0	25.0	
Q2733-01DUP	TW-WTS-12DUP	DUP	WATER	08/01/2025	50.0	25.0	
Q2733-01MS	TW-WTS-12MS	MS	WATER	08/01/2025	50.0	25.0	
Q2733-01MSD	TW-WTS-12MSD	MSD	WATER	08/01/2025	50.0	25.0	
Q2742-01	9 Banker	SAM	WATER	08/01/2025	50.0	25.0	

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136697

Review By	Janvi	Review On	8/5/2025 10:04:40 AM
Supervise By	jaswal	Supervise On	8/6/2025 7:45:51 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/04/25 11:15		Jaswal	OK
2	S1	S1	CAL2	08/04/25 11:20		Jaswal	OK
3	S2	S2	CAL3	08/04/25 11:24		Jaswal	OK
4	S3	S3	CAL4	08/04/25 11:28		Jaswal	OK
5	S4	S4	CAL5	08/04/25 11:32		Jaswal	OK
6	S5	S5	CAL6	08/04/25 11:36		Jaswal	OK
7	ICV01	ICV01	ICV	08/04/25 11:41		Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/04/25 11:45		Jaswal	OK
9	ICB01	ICB01	ICB	08/04/25 11:52		Jaswal	OK
10	CRI01	CRI01	CRDL	08/04/25 11:57		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/04/25 12:01		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/04/25 12:05		Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/04/25 12:09		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/04/25 12:14		Jaswal	OK
15	CCV01	CCV01	CCV	08/04/25 12:25		Jaswal	OK
16	CCB01	CCB01	CCB	08/04/25 12:29		Jaswal	OK
17	PB169076BL	PB169076BL	MB	08/04/25 12:55		Jaswal	OK
18	PB169076BS	PB169076BS	LCS	08/04/25 13:00		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136697

Review By	Janvi	Review On	8/5/2025 10:04:40 AM
Supervise By	jaswal	Supervise On	8/6/2025 7:45:51 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

19	Q2725-02DL	CompDL	SAM	08/04/25 13:04	Straight 5x dilution for all elements	Jaswal	OK
20	Q2725-02DUPDL	CompDUPDL	DUP	08/04/25 13:08	Straight 5x dilution for all elements	Jaswal	OK
21	Q2725-02LDL	CompLDL	SD	08/04/25 13:12	Straight 25x dilution for all elements	Jaswal	OK
22	Q2725-02MSDL	CompMSDL	MS	08/04/25 13:16	Straight 5x dilution for all elements	Jaswal	OK
23	Q2725-02MSDDL	CompMSDDL	MSD	08/04/25 13:20	Straight 5x dilution for all elements	Jaswal	OK
24	Q2725-02ADL	CompADL	PS	08/04/25 13:24	Straight 5x dilution for all elements (0.1ml m6004,m6013-10ml sample before dilution)	Jaswal	OK
25	PB169078BL	PB169078BL	MB	08/04/25 13:29		Jaswal	OK
26	PB169078BS	PB169078BS	LCS	08/04/25 13:33		Jaswal	OK
27	CCV02	CCV02	CCV	08/04/25 13:37		Jaswal	OK
28	CCB02	CCB02	CCB	08/04/25 13:41		Jaswal	OK
29	Q2733-01	TW-WTS-12	SAM	08/04/25 13:45		Jaswal	OK
30	Q2733-01DUP	TW-WTS-12DUP	DUP	08/04/25 13:50		Jaswal	OK
31	Q2733-01L	TW-WTS-12L	SD	08/04/25 13:54		Jaswal	OK
32	Q2733-01MS	TW-WTS-12MS	MS	08/04/25 13:59		Jaswal	OK
33	Q2733-01MSD	TW-WTS-12MSD	MSD	08/04/25 14:03		Jaswal	OK
34	Q2733-01A	TW-WTS-12A	PS	08/04/25 14:07	0.1ml m6004,m6013-10ml sample	Jaswal	OK
35	Q2742-01	9 Banker	SAM	08/04/25 14:12		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136697

Review By	Janvi	Review On	8/5/2025 10:04:40 AM
Supervise By	jaswal	Supervise On	8/6/2025 7:45:51 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

36	Q2734-01	VNJ-253	SAM	08/04/25 14:17		Jaswal	OK
37	Q2741-01	TR-05-07312025	SAM	08/04/25 14:21		Jaswal	OK
38	Q2743-01	WC1	SAM	08/04/25 14:25		Jaswal	OK
39	CCV03	CCV03	CCV	08/04/25 14:43		Jaswal	OK
40	CCB03	CCB03	CCB	08/04/25 14:54		Jaswal	OK
41	Q2753-01	289	SAM	08/04/25 14:58		Jaswal	OK
42	Q2758-01	WASTE	SAM	08/04/25 15:03	%soil missing	Jaswal	OK
43	PB169096BL	PB169096BL	MB	08/04/25 15:07		Jaswal	OK
44	Q2740-01	OR-02-07312025	SAM	08/04/25 15:15		Jaswal	OK
45	Q2740-01DUP	OR-02-07312025DUP	DUP	08/04/25 15:19		Jaswal	OK
46	Q2740-01L	OR-02-07312025L	SD	08/04/25 15:23		Jaswal	OK
47	Q2740-01MS	OR-02-07312025MS	MS	08/04/25 15:27		Jaswal	OK
48	Q2740-01MSD	OR-02-07312025MSD	MSD	08/04/25 15:31		Jaswal	OK
49	Q2740-01A	OR-02-07312025A	PS	08/04/25 15:35	0.1ml m6004,m6013-10ml sample	Jaswal	OK
50	PB169096BS	PB169096BS	LCS	08/04/25 15:39		Jaswal	OK
51	CCV04	CCV04	CCV	08/04/25 15:48		Jaswal	OK
52	CCB04	CCB04	CCB	08/04/25 15:53		Jaswal	OK
53	Q2734-02	VNJ-253	SAM	08/04/25 16:08		Jaswal	OK
54	Q2703-05	TP-4	SAM	08/04/25 16:12		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136697

Review By	Janvi	Review On	8/5/2025 10:04:40 AM
Supervise By	jaswal	Supervise On	8/6/2025 7:45:51 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

55	Q2743-02	WC1	SAM	08/04/25 16:16		Jaswal	OK
56	Q2753-02	289	SAM	08/04/25 16:21		Jaswal	OK
57	Q2753-02DUP	289DUP	DUP	08/04/25 16:25		Jaswal	OK
58	Q2753-02L	289L	SD	08/04/25 16:30		Jaswal	OK
59	Q2753-02MS	289MS	MS	08/04/25 16:34		Jaswal	OK
60	Q2753-02MSD	289MSD	MSD	08/04/25 16:38		Jaswal	OK
61	Q2753-02A	289A	PS	08/04/25 16:42	0.1ml m6004,m6013-10ml sample	Jaswal	OK
62	PB169079TB	PB169079TB	MB	08/04/25 16:46		Jaswal	OK
63	CCV05	CCV05	CCV	08/04/25 16:51		Jaswal	OK
64	CCB05	CCB05	CCB	08/04/25 16:55		Jaswal	OK
65	PB169101BL	PB169101BL	MB	08/04/25 16:59		Jaswal	OK
66	PB169101BS	PB169101BS	LCS	08/04/25 17:04		Jaswal	OK
67	Q2745-01	RW8-SP100-2025073	SAM	08/04/25 17:08		Jaswal	OK
68	Q2745-02	RW8-SP303-2025073	SAM	08/04/25 17:12		Jaswal	OK
69	Q2756-01	71725	SAM	08/04/25 17:16	CCV fail for TI	Jaswal	Not Ok
70	Q2756-01DUP	71725DUP	DUP	08/04/25 17:20	CCV fail for TI	Jaswal	Not Ok
71	Q2756-01L	71725L	SD	08/04/25 17:25	CCV fail for TI	Jaswal	Not Ok
72	Q2756-01MS	71725MS	MS	08/04/25 17:29	CCV fail for TI	Jaswal	Not Ok
73	Q2756-01MSD	71725MSD	MSD	08/04/25 17:33	CCV fail for TI	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136697

Review By	Janvi	Review On	8/5/2025 10:04:40 AM
Supervise By	jaswal	Supervise On	8/6/2025 7:45:51 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86452,MP86453,MP86454,MP86455,MP86456,MP86458		
ICV Standard	MP86459		
CCV Standard	MP86462		
ICSA Standard	MP86460,MP86461		
CRI Standard	MP86458		
LCS Standard			
Chk Standard	MP86463,MP86464		

74	Q2756-01A	71725A	PS	08/04/25 17:37	CCV fail for TI	Jaswal	Not Ok
75	CCV06	CCV06	CCV	08/04/25 17:41	CCV fail for TI	Jaswal	OK
76	CCB06	CCB06	CCB	08/04/25 17:45		Jaswal	OK
77	PB169098BL	PB169098BL	MB	08/04/25 17:49	CCV fail for TI	Jaswal	Not Ok
78	PB169098BS	PB169098BS	LCS	08/04/25 17:54	CCV fail for TI	Jaswal	Not Ok
79	CCV07	CCV07	CCV	08/04/25 17:58		Jaswal	OK
80	CCB07	CCB07	CCB	08/04/25 18:02		Jaswal	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BLOCK #1 2. N/A

Start Digest Date: 08/01/2025 Time : 10:05 Temp : 96 °C

End Digest Date: 08/01/2025 Time : 13:10 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: *SLB*

Supervisor Signature: *[Signature]*

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6007
LFS-2	0.25	M6015
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP85156
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK# CELL#50 96C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/01/25 14:10	SLB met dig	<i>[Signature]</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169078BL	PBW078	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	15
PB169078BS	LCS078	<2	50	25	Colorless	Colorless	Clear	Clear	M6007,M6015	16
Q2733-01MS	TW-WTS-12MS	<2	50	25	Colorless	Colorless	Clear	Clear	M6007,M6015	19
Q2733-01MSD	TW-WTS-12MSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6007,M6015	20
Q2733-01DUP	TW-WTS-12DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	18
Q2733-01	TW-WTS-12	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	17
Q2742-01	9 BANKER	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	21



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	07/31/25 12:30
Project:	Banker	Date Received:	07/31/25
Client Sample ID:	9 Banker	SDG No.:	Q2742
Lab Sample ID:	Q2742-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Chloride	914	OR	1	0.19	0.60	mg/L		08/01/25 12:26	300.0

Comments: _____

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	07/31/25 12:30
Project:	Banker	Date Received:	07/31/25
Client Sample ID:	9 BankerDL	SDG No.:	Q2742
Lab Sample ID:	Q2742-01DL	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Chloride	590	D	200	38.0	120	mg/L		08/01/25 13:31	300.0

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements
 H = Sample Analysis Out Of Hold Time

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits



QC RESULT SUMMARY

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2742

Project: Banker

RunNo.: LB136667

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.5	10	95	90-110	07/21/2025
Chloride	mg/L	2.8	3	93	90-110	07/21/2025
Fluoride	mg/L	1.9	2	95	90-110	07/21/2025
Nitrite	mg/L	2.8	3	93	90-110	07/21/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/21/2025
Sulfate	mg/L	14.2	15	95	90-110	07/21/2025
Orthophosphate as P	mg/L	4.8	5	96	90-110	07/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.2	10	102	90-110	08/01/2025
Chloride	mg/L	3.1	3	103	90-110	08/01/2025
Fluoride	mg/L	2.1	2	105	90-110	08/01/2025
Nitrite	mg/L	3.1	3	103	90-110	08/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/01/2025
Sulfate	mg/L	15.3	15	102	90-110	08/01/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	08/01/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	08/01/2025
Chloride	mg/L	3.1	3	103	90-110	08/01/2025
Fluoride	mg/L	2	2	100	90-110	08/01/2025
Nitrite	mg/L	3.1	3	103	90-110	08/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/01/2025
Sulfate	mg/L	15.2	15	101	90-110	08/01/2025
Orthophosphate as P	mg/L	4.6	5	92	90-110	08/01/2025

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q2742

Project: Banker

RunNo.: LB136667

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	07/21/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	07/21/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	07/21/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	07/21/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	07/21/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	07/21/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	07/21/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	08/01/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	08/01/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	08/01/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	08/01/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	08/01/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	08/01/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	08/01/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	08/01/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	08/01/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	08/01/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	08/01/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	08/01/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	08/01/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	08/01/2025

Preparation Blank Summary

Client: G Environmental

SDG No.: Q2742

Project: Banker

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136667BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	08/01/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	08/01/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	08/01/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	08/01/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	08/01/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	08/01/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	08/01/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2742
Project:	Banker	Sample ID:	Q2742-01
Client ID:	9 BankerMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.2		0.37	U	10	1	102		08/01/2025
Chloride	mg/L	80-120	872	OR	914	OR	3	1	-1400	*	08/01/2025
Fluoride	mg/L	80-120	2.20		0.16	J	2	1	102		08/01/2025
Nitrite	mg/L	80-120	3.00		0.083	J	3	1	97		08/01/2025
Nitrate	mg/L	80-120	7.10	OR	4.50		2.5	1	104		08/01/2025
Sulfate	mg/L	80-120	45.7	OR	30.9		15	1	99		08/01/2025
Orthophosphate as P	mg/L	80-120	4.50		0.34	U	5	1	90		08/01/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q2742
Project:	Banker	Sample ID:	Q2742-01
Client ID:	9 BankerMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.1		0.37	U	10	1	101		08/01/2025
Chloride	mg/L	80-120	872	OR	914	OR	3	1	-1400	*	08/01/2025
Fluoride	mg/L	80-120	2.10		0.16	J	2	1	97		08/01/2025
Nitrite	mg/L	80-120	3.00		0.083	J	3	1	97		08/01/2025
Nitrate	mg/L	80-120	6.90	OR	4.50		2.5	1	96		08/01/2025
Sulfate	mg/L	80-120	45.3	OR	30.9		15	1	96		08/01/2025
Orthophosphate as P	mg/L	80-120	4.80		0.34	U	5	1	96		08/01/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q2742
Project:	Banker	Sample ID:	Q2742-01
Client ID:	9 BankerMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Nitrite	mg/L	+/-20	3.00		3.00		1	0		08/01/2025
Chloride	mg/L	+/-20	872	OR	872	OR	1	0		08/01/2025
Bromide	mg/L	+/-20	10.2		10.1		1	1		08/01/2025
Sulfate	mg/L	+/-20	45.7	OR	45.3	OR	1	1		08/01/2025
Nitrate	mg/L	+/-20	7.10	OR	6.90	OR	1	3		08/01/2025
Fluoride	mg/L	+/-20	2.20		2.10		1	5		08/01/2025
Orthophosphate as P	mg/L	+/-20	4.50		4.80		1	6		08/01/2025

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2742

Project: Banker

Run No.: LB136667

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136667BSW							
Bromide	mg/L	10	10.3		103	1	90-110	08/01/2025
Chloride	mg/L	3	3.10		103	1	90-110	08/01/2025
Fluoride	mg/L	2	2.10		105	1	90-110	08/01/2025
Nitrite	mg/L	3	3.10		103	1	90-110	08/01/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	08/01/2025
Sulfate	mg/L	15	15.4		103	1	90-110	08/01/2025
Orthophosphate as P	mg/L	5	4.90		98	1	90-110	08/01/2025

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136667

Review By	rubina	Review On	8/5/2025 9:47:33 AM
Supervise By	Iwona	Supervise On	8/5/2025 9:49:15 AM
SubDirectory	LB136667	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114024,WP114025,WP114026,WP114027,WP114028,WP114029,WP114030		
ICV Standard	WP114031		
CCV Standard	WP116146		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114147		
Chk Standard	WP114032,WP114033		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	07/21/25 15:51	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	07/21/25 16:12	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	07/21/25 16:34	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	07/21/25 16:55		RM/IZ	OK
5	STD5	STD5	CAL5	07/21/25 17:17		RM/IZ	OK
6	STD6	STD6	CAL6	07/21/25 17:38		RM/IZ	OK
7	STD7	STD7	CAL7	07/21/25 18:00		RM/IZ	OK
8	ICV1	ICV1	ICV	07/21/25 18:21		RM/IZ	OK
9	ICB1	ICB1	ICB	07/21/25 18:42		RM/IZ	OK
10	CCV1	CCV1	CCV	08/01/25 11:00		RM/IZ	OK
11	CCB1	CCB1	CCB	08/01/25 11:22		RM/IZ	OK
12	LB136667BLW	LB136667BLW	MB	08/01/25 11:43		RM/IZ	OK
13	LB136667BSW	LB136667BSW	LCS	08/01/25 12:05		RM/IZ	OK
14	Q2742-01	9 Banker	SAM	08/01/25 12:26	Cl is High	RM/IZ	Dilution
15	Q2742-01MS	9 BankerMS	MS	08/01/25 12:48	9.5ml of sample, 0.5mL W3092	RM/IZ	OK
16	Q2742-01MSD	9 BankerMSD	MSD	08/01/25 13:09	9.5ml of sample, 0.5mL W3092	RM/IZ	OK
17	Q2742-01DL	9 BankerDL	SAM	08/01/25 13:31	200Xfor Cl	RM/IZ	Confirms
18	CCV2	CCV2	CCV	08/01/25 13:53		RM/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136667

Review By	rubina	Review On	8/5/2025 9:47:33 AM
Supervise By	Iwona	Supervise On	8/5/2025 9:49:15 AM
SubDirectory	LB136667	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114024,WP114025,WP114026,WP114027,WP114028,WP114029,WP114030		
ICV Standard	WP114031		
CCV Standard	WP116146		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114147		
Chk Standard	WP114032,WP114033		

19	CCB2	CCB2	CCB	08/01/25 14:14		RM/IZ	OK
----	------	------	-----	----------------	--	-------	----

LAB CHRONICLE

OrderID:	Q2742	OrderDate:	7/31/2025 3:56:00 PM
Client:	G Environmental	Project:	Banker
Contact:	Gary Landis	Location:	O22

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2742-01	9 Banker	WATER			07/31/25 12:30			07/31/25
			Anions Group1	300.0			08/01/25 12:26	
Q2742-01DL	9 BankerDL	WATER			07/31/25 12:30			07/31/25
			Anions Group1	300.0			08/01/25 13:31	



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:
COMPANY: Geop Inc
ADDRESS: 8 CARRIAGE
CITY: Succasunna STATE: NJ ZIP: 07876
ATTENTION:
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Banker
PROJECT NO.: LOCATION:
PROJECT MANAGER: GL
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Geop Inc PO#: 8 CARRIAGE
ADDRESS: 8 CARRIAGE
CITY: Succasunna STATE: NJ ZIP: 07876
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) 5-DAY DAYS*
HARDCOPY (DATA PACKAGE) DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
EDD FORMAT: waste, Njdep

1	2	3	4	5	6	7	8	9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										← Specify Preservatives		
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	A-HCl B-HNO3 C-H2SO4	D-NaOH E-ICE F-OTHER	
1.	9 Banker	DW	X		7/3/25	1230	2	X		X									
2.																			
3.																			
4.																			
5.																			
6.																			
7.																			
8.																			
9.																			
10.																			

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. [Signature]	7/3/25	1. [Signature] 15:52	Comments: 9 Banker 39
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3.		3.	

Page ____ of ____

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488