

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136667

Review By	rubina	Review On	8/5/2025 9:47:33 AM
Supervise By	Iwona	Supervise On	8/5/2025 9:49:15 AM
SubDirectory	LB136667	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114024,WP114025,WP114026,WP114027,WP114028,WP114029,WP114030		
ICV Standard	WP114031		
CCV Standard	WP116146		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114147		
Chk Standard	WP114032,WP114033		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	07/21/25 15:51	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	07/21/25 16:12	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	07/21/25 16:34	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	07/21/25 16:55		RM/IZ	OK
5	STD5	STD5	CAL5	07/21/25 17:17		RM/IZ	OK
6	STD6	STD6	CAL6	07/21/25 17:38		RM/IZ	OK
7	STD7	STD7	CAL7	07/21/25 18:00		RM/IZ	OK
8	ICV1	ICV1	ICV	07/21/25 18:21		RM/IZ	OK
9	ICB1	ICB1	ICB	07/21/25 18:42		RM/IZ	OK
10	CCV1	CCV1	CCV	08/01/25 11:00		RM/IZ	OK
11	CCB1	CCB1	CCB	08/01/25 11:22		RM/IZ	OK
12	LB136667BLW	LB136667BLW	MB	08/01/25 11:43		RM/IZ	OK
13	LB136667BSW	LB136667BSW	LCS	08/01/25 12:05		RM/IZ	OK
14	Q2742-01	9 Banker	SAM	08/01/25 12:26	Cl is High	RM/IZ	Dilution
15	Q2742-01MS	9 BankerMS	MS	08/01/25 12:48	9.5ml of sample, 0.5mL W3092	RM/IZ	OK
16	Q2742-01MSD	9 BankerMSD	MSD	08/01/25 13:09	9.5ml of sample, 0.5mL W3092	RM/IZ	OK
17	Q2742-01DL	9 BankerDL	SAM	08/01/25 13:31	200Xfor Cl	RM/IZ	Confirms
18	CCV2	CCV2	CCV	08/01/25 13:53		RM/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136667

Review By	rubina	Review On	8/5/2025 9:47:33 AM
Supervise By	Iwona	Supervise On	8/5/2025 9:49:15 AM
SubDirectory	LB136667	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114024,WP114025,WP114026,WP114027,WP114028,WP114029,WP114030		
ICV Standard	WP114031		
CCV Standard	WP116146		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP114147		
Chk Standard	WP114032,WP114033		

19	CCB2	CCB2	CCB	08/01/25 14:14		RM/IZ	OK
----	------	------	-----	----------------	--	-------	----