

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2742

Project: Banker

RunNo.: LB136667

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	9.5	10	95	90-110	07/21/2025
Chloride	mg/L	2.8	3	93	90-110	07/21/2025
Fluoride	mg/L	1.9	2	95	90-110	07/21/2025
Nitrite	mg/L	2.8	3	93	90-110	07/21/2025
Nitrate	mg/L	2.3	2.5	92	90-110	07/21/2025
Sulfate	mg/L	14.2	15	95	90-110	07/21/2025
Orthophosphate as P	mg/L	4.8	5	96	90-110	07/21/2025
Sample ID: CCV1						
Bromide	mg/L	10.2	10	102	90-110	08/01/2025
Chloride	mg/L	3.1	3	103	90-110	08/01/2025
Fluoride	mg/L	2.1	2	105	90-110	08/01/2025
Nitrite	mg/L	3.1	3	103	90-110	08/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/01/2025
Sulfate	mg/L	15.3	15	102	90-110	08/01/2025
Orthophosphate as P	mg/L	4.9	5	98	90-110	08/01/2025
Sample ID: CCV2						
Bromide	mg/L	10.2	10	102	90-110	08/01/2025
Chloride	mg/L	3.1	3	103	90-110	08/01/2025
Fluoride	mg/L	2	2	100	90-110	08/01/2025
Nitrite	mg/L	3.1	3	103	90-110	08/01/2025
Nitrate	mg/L	2.5	2.5	100	90-110	08/01/2025
Sulfate	mg/L	15.2	15	101	90-110	08/01/2025
Orthophosphate as P	mg/L	4.6	5	92	90-110	08/01/2025