

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q2742

Project: Banker

Run No.: LB136667

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136667BSW							
Bromide	mg/L	10	10.3		103	1	90-110	08/01/2025
Chloride	mg/L	3	3.10		103	1	90-110	08/01/2025
Fluoride	mg/L	2	2.10		105	1	90-110	08/01/2025
Nitrite	mg/L	3	3.10		103	1	90-110	08/01/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	08/01/2025
Sulfate	mg/L	15	15.4		103	1	90-110	08/01/2025
Orthophosphate as P	mg/L	5	4.90		98	1	90-110	08/01/2025