

Analytical Summary Report

Analysis Method: 9045D

Analyst By : Eman

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136674

Slope : 99.4

BalanceID: WC SC-7

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	08/01/2025	11:00
2	CAL2	1	Water	NA	NA	20.3	7.01	08/01/2025	11:03
3	CAL3	1	Water	NA	NA	20.3	9.99	08/01/2025	11:05
4	ICV	1	Water	NA	NA	20.3	7.00	08/01/2025	11:08
5	CCV1	1	Water	NA	NA	20.4	2.01	08/01/2025	11:50
6	Q2743-01	1	Solid	20	20	24.4	7.75	08/01/2025	11:54
7	Q2743-01DUP	1	Solid	20	20	24.4	7.74	08/01/2025	11:58
8	CCV2	1	Water	NA	NA	20.4	12.02	08/01/2025	12:03

WORKLIST(Hardcopy Internal Chain)

LB 136674

WorkList Name : PH-080125-2

WorkList ID : 191060

Department : Wet-Chemistry

Date : 08-01-2025 09:05:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2743-01	WC1	Solid	pH	Cool 4 deg C	GENV01	O13	07/31/2025	9045D

Date/Time 08/01/25 10:50
 Raw Sample Received by: EM (wc)
 Raw Sample Relinquished by: CP SM

Date/Time 08/01/25 12:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: EM (wc)