

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2743

Project: Capra

RunNo.: LB136670

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	08/01/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	08/01/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	08/01/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	08/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2743

Project: Capra

RunNo.: LB136674

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	08/01/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	08/01/2025
Sample ID: CCV2 pH	pH	12.02	12	100	90-110	08/01/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q2743

Project: Capra

RunNo.: LB136676

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Corrosivity		pH	7.00	7	100	90-110	08/01/2025
Sample ID:	CCV1						
Corrosivity		pH	2.00	2.00	100	90-110	08/01/2025
Sample ID:	CCV2						
Corrosivity		pH	12.02	12.00	100	90-110	08/01/2025