

LB136668


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

46960



A= 0.0000e+000

B= 2.1292e-004

C= -2.4451e-002

Rho= 0.9987914

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.097	-0.097	-340	0.000	-340					-
0.2	0.200	0.196	-0.004	1036	0.0 %	1036					-2
2.5	2.500	2.818	0.318	13348	0.0 %	13348					13
5.0	5.000	4.720	-0.280	22282	0.0 %	22282					-5
7.5	7.500	7.589	0.089	35758	0.0 %	35758					3
10.0	10.000	9.974	-0.026	46960	0.0 %	46960					0

LB136668 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Aug 2025 11:09:40

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	5.0	-	-340	PPB	Std	0.0000	7471B	01 Aug 2025 11:48:28
S	0.2 - 1	5.2	-	1036	PPB	Std	0.2000	7471B	01 Aug 2025 11:56:00
S	2.5 - 1	5.2.5	-	13348	PPB	Std	2.5000	7471B	01 Aug 2025 12:05:23
S	5.0 - 1	5.5	-	22282	PPB	Std	5.0000	7471B	01 Aug 2025 12:10:38
S	7.5 - 1	5.7.5	-	35758	PPB	Std	7.5000	7471B	01 Aug 2025 12:12:54
S	10.0 - 1	5.16	-	46960	PPB	Std	10.0000	7471B	01 Aug 2025 12:19:06
U	ICV48 - 1	ICV48	3.8760	18319	PPB	SMPL	-7471B		01 Aug 2025 12:22:51
U	ICB48 - 1	ICB48	-0.0268	-11	PPB	SMPL	-7471B		01 Aug 2025 12:42:48
U	CCV73 - 1	CCV73	4.8940	23100	PPB	SMPL	-7471B		01 Aug 2025 12:45:03
U	CCB73 - 1	CCB73	-0.0736	-231	PPB	SMPL	-7471B		01 Aug 2025 12:47:18
U	CRA - 1	CRA	0.2183	1140	PPB	SMPL	-7471B		01 Aug 2025 12:49:37
U	HighStd - 1	HighStd	10.1333	47707	PPB	SMPL	-7471B		01 Aug 2025 12:51:52
U	ChkStd - 1	ChkStd	7.5836	35732	PPB	SMPL	-7471B		01 Aug 2025 12:54:07
U	PB169081BL - 1	PBS	-0.0342	-46	PPB	SMPL	-7471B		01 Aug 2025 12:59:23
U	PB169081BS - 1	LCSS	4.1345	19533	PPB	SMPL	-7471B		01 Aug 2025 13:01:38
U	Q2726-01 - 1	WBR-PBO	2.4441	11594	PPB	SMPL	-7471B		01 Aug 2025 13:03:55
U	Q2728-01 - 1	7211931-VNJ240	2.2408	10639	PPB	SMPL	-7471B		01 Aug 2025 13:06:15
U	Q2728-01DUP - 1	7211931-VNJ240DUP	2.4020	11396	PPB	SMPL	-7471B		01 Aug 2025 13:08:36
U	Q2728-01MS - 1	7211931-VNJ240MS	6.0425	28494	PPB	SMPL	-7471B		01 Aug 2025 13:10:56
U	Q2728-01MSD - 1	7211931-VNJ240MSD	5.8911	27783	PPB	SMPL	-7471B		01 Aug 2025 13:13:14
U	CCV74 - 1	CCV74	4.6791	22091	PPB	SMPL	-7471B		01 Aug 2025 13:15:36
U	CCB74 - 1	CCB74	-0.0074	80	PPB	SMPL	-7471B		01 Aug 2025 13:22:05
U	Q2743-01 - 1	WC1	46.7076	219483	PPB	SMPL	-7471B		01 Aug 2025 13:24:20
U	Q2728-01LX5 - 1		0.0973	572	PPB	SMPL	-7471B		01 Aug 2025 13:26:35
U	Q2728-01A - 1		7.1269	33587	PPB	SMPL	-7471B		01 Aug 2025 13:29:36
U	Q2743-01DLX10 - 1		4.0078	18938	PPB	SMPL	-7471B		01 Aug 2025 13:44:41
U	CCV75 - 1	CCV75	4.6817	22103	PPB	SMPL	-7471B		01 Aug 2025 13:46:57
U	CCB75 - 1	CCB75	-0.0455	-99	PPB	SMPL	-7471B		01 Aug 2025 13:49:17