



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	109	70 - 130	CV	08/01/2025	12:49	LB136668
CRI01	Aluminum	111	100	112	65 - 135	P	08/04/2025	11:57	LB136697
	Antimony	52.6	50.0	105	65 - 135	P	08/04/2025	11:57	LB136697
	Arsenic	20.1	20.0	101	65 - 135	P	08/04/2025	11:57	LB136697
	Barium	103	100	103	65 - 135	P	08/04/2025	11:57	LB136697
	Beryllium	6.41	6.0	107	65 - 135	P	08/04/2025	11:57	LB136697
	Cadmium	6.14	6.0	102	65 - 135	P	08/04/2025	11:57	LB136697
	Calcium	2080	2000	104	65 - 135	P	08/04/2025	11:57	LB136697
	Chromium	9.85	10.0	98	65 - 135	P	08/04/2025	11:57	LB136697
	Cobalt	31.2	30.0	104	65 - 135	P	08/04/2025	11:57	LB136697
	Copper	22.9	20.0	115	65 - 135	P	08/04/2025	11:57	LB136697
	Iron	111	100	111	65 - 135	P	08/04/2025	11:57	LB136697
	Lead	11.2	12.0	93	65 - 135	P	08/04/2025	11:57	LB136697
	Magnesium	2210	2000	110	65 - 135	P	08/04/2025	11:57	LB136697
	Manganese	20.5	20.0	103	65 - 135	P	08/04/2025	11:57	LB136697
	Nickel	41.8	40.0	105	65 - 135	P	08/04/2025	11:57	LB136697
	Potassium	2030	2000	101	65 - 135	P	08/04/2025	11:57	LB136697
	Selenium	22.0	20.0	110	65 - 135	P	08/04/2025	11:57	LB136697
	Silver	11.1	10.0	110	65 - 135	P	08/04/2025	11:57	LB136697
	Sodium	2030	2000	102	65 - 135	P	08/04/2025	11:57	LB136697
	Thallium	46.4	40.0	116	65 - 135	P	08/04/2025	11:57	LB136697
	Vanadium	43.2	40.0	108	65 - 135	P	08/04/2025	11:57	LB136697
	Zinc	44.8	40.0	112	65 - 135	P	08/04/2025	11:57	LB136697