

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169081BS Mercury	mg/Kg	0.26	0.27		103	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169096BS							
Aluminum	mg/Kg	100	104		104	80 - 120	P
Antimony	mg/Kg	40.0	44.3		111	80 - 120	P
Arsenic	mg/Kg	40.0	42.0		105	80 - 120	P
Barium	mg/Kg	10.0	10.7		107	80 - 120	P
Beryllium	mg/Kg	10.0	10.9		110	80 - 120	P
Cadmium	mg/Kg	10.0	9.72		97	80 - 120	P
Calcium	mg/Kg	50.0	55.5	J	111	80 - 120	P
Chromium	mg/Kg	20.0	19.7		98	80 - 120	P
Cobalt	mg/Kg	10.0	10.1		101	80 - 120	P
Copper	mg/Kg	15.0	16.2		108	80 - 120	P
Iron	mg/Kg	150	159		106	80 - 120	P
Lead	mg/Kg	50.0	49.4		99	80 - 120	P
Magnesium	mg/Kg	100	103		103	80 - 120	P
Manganese	mg/Kg	10.0	10.6		106	80 - 120	P
Nickel	mg/Kg	25.0	25.4		102	80 - 120	P
Potassium	mg/Kg	500	523		105	80 - 120	P
Selenium	mg/Kg	100	110		110	80 - 120	P
Silver	mg/Kg	3.8	4.05		107	80 - 120	P
Sodium	mg/Kg	150	148		99	80 - 120	P
Thallium	mg/Kg	100	113		113	80 - 120	P
Vanadium	mg/Kg	15.0	15.3		102	80 - 120	P
Zinc	mg/Kg	10.0	11.1		111	80 - 120	P