


7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.026	0.026	200	0.000	200	0				31
0.2	0.200	0.222	0.022	957	0.0 %	957					2
2.5	2.500	2.549	0.049	9928	0.0 %	9928	0				-2
5.0	5.000	4.877	-0.123	18902	0.0 %	18902					-1
7.5	7.500	7.424	-0.076	28722	0.0 %	28722					1
10.0	10.000	10.102	0.102	39044	0.0 %	39044					

LB136721 INSTRUMENT ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Aug 2025 08:46:28

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std CondMethod	Date
S	0.0 - 1	50	-	200	PPB	Std	0.00007470A	06 Aug 2025 08:55:59
S	0.2 - 1	50.2	-	957	PPB	Std	0.20007470A	06 Aug 2025 08:58:16
S	2.5 - 1	52.5	-	9928	PPB	Std	2.50007470A	06 Aug 2025 09:00:33
S	5.0 - 1	55.0	-	18902	PPB	Std	5.00007470A	06 Aug 2025 09:02:50
S	7.5 - 1	57.5	-	28722	PPB	Std	7.50007470A	06 Aug 2025 09:05:06
S	10.0 - 1	510.0	-	39044	PPB	Std	10.00007470A	06 Aug 2025 09:07:24
U	ICV54 - 1	ICV54	4.0897	15867	PPB	SMPL	-7470A	06 Aug 2025 09:10:12
U	ICB54 - 1	ICB54	0.0310	221	PPB	SMPL	-7470A	06 Aug 2025 09:12:29
U	CCV89 - 1	CCV89	4.8285	18715	PPB	SMPL	-7470A	06 Aug 2025 09:14:46
U	CCB89 - 1	CCB89	0.0352	237	PPB	SMPL	-7470A	06 Aug 2025 09:17:02
U	CRA - 1	CRA	0.2378	1018	PPB	SMPL	-7470A	06 Aug 2025 09:19:18
U	HighStd - 1	HighStd	10.0841	38975	PPB	SMPL	-7470A	06 Aug 2025 09:21:33
U	ChkStd - 1	ChkStd	7.4363	28768	PPB	SMPL	-7470A	06 Aug 2025 09:26:37
U	PB169140BL - 1	PBW	0.0702	372	PPB	SMPL	-7470A	06 Aug 2025 09:28:53
U	PB169140BS - 1	LCSW	4.0799	15829	PPB	SMPL	-7470A	06 Aug 2025 09:31:13
U	Q2734-02 - 1	VNJ-253	0.0852	430	PPB	SMPL	-7470A	06 Aug 2025 09:33:30
U	Q2743-02 - 1	WC1	0.1317	609	PPB	SMPL	-7470A	06 Aug 2025 09:35:48
U	Q2753-02 - 1	289	0.0186	173	PPB	SMPL	-7470A	06 Aug 2025 09:38:05
U	Q2753-02DUP - 1	289DUP	0.0487	289	PPB	SMPL	-7470A	06 Aug 2025 09:40:22
U	Q2753-02MS - 1	289MS	3.9937	15497	PPB	SMPL	-7470A	06 Aug 2025 09:45:16
U	CCV90 - 1	CCV90	4.8173	18672	PPB	SMPL	-7470A	06 Aug 2025 09:49:01
U	CCB90 - 1	CCB90	0.0676	362	PPB	SMPL	-7470A	06 Aug 2025 09:51:17
U	Q2753-02MSD - 1	289MSD	3.8272	14855	PPB	SMPL	-7470A	06 Aug 2025 09:53:34
U	PB169079TB - 1	PB169079TB	0.0720	379	PPB	SMPL	-7470A	06 Aug 2025 09:55:49
U	Q2753-02LX5 - 1		-0.0014	96	PPB	SMPL	-7470A	06 Aug 2025 09:58:06
U	Q2753-02A - 1		4.0640	15768	PPB	SMPL	-7470A	06 Aug 2025 10:00:22
U	CCV91 - 1	CCV91	4.7756	18511	PPB	SMPL	-7470A	06 Aug 2025 10:02:38
U	CCB91 - 1	CCB91	0.0440	271	PPB	SMPL	-7470A	06 Aug 2025 10:04:56