



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	20.1	20.0	101	65 - 135	P	08/04/2025	11:57	LB136697
	Barium	103	100	103	65 - 135	P	08/04/2025	11:57	LB136697
	Cadmium	6.14	6.0	102	65 - 135	P	08/04/2025	11:57	LB136697
	Chromium	9.85	10.0	98	65 - 135	P	08/04/2025	11:57	LB136697
	Lead	11.2	12.0	93	65 - 135	P	08/04/2025	11:57	LB136697
	Selenium	22.0	20.0	110	65 - 135	P	08/04/2025	11:57	LB136697
	Silver	11.1	10.0	110	65 - 135	P	08/04/2025	11:57	LB136697
CRA	Mercury	0.24	0.2	119	70 - 130	CV	08/06/2025	09:19	LB136721