

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>G Environmental</u>	SDG No.: <u>Q2743</u>
Contract: <u>GENV01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	9.45			-20	20	08/04/2025	12:01	LB136697
	Barium	6.24	6.0	104	-94	106	08/04/2025	12:01	LB136697
	Cadmium	-0.56	1.0	56	-5	7	08/04/2025	12:01	LB136697
	Chromium	56.1	52.0	108	42	62	08/04/2025	12:01	LB136697
	Lead	-8.46			-12	12	08/04/2025	12:01	LB136697
	Selenium	0.23			-20	20	08/04/2025	12:01	LB136697
	Silver	2.22			-10	10	08/04/2025	12:01	LB136697
ICSAB01	Arsenic	106	104	102	88.4	120	08/04/2025	12:05	LB136697
	Barium	510	537	95	437	637	08/04/2025	12:05	LB136697
	Cadmium	983	972	101	826	1120	08/04/2025	12:05	LB136697
	Chromium	553	542	102	460	624	08/04/2025	12:05	LB136697
	Lead	37.5	49.0	76	37	61	08/04/2025	12:05	LB136697
	Selenium	42.1	46.0	92	26	66	08/04/2025	12:05	LB136697
	Silver	214	201	106	170	232	08/04/2025	12:05	LB136697