

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169101BS							
Arsenic	ug/L	4000	4170		104	80 - 120	P
Barium	ug/L	1000	1100		110	80 - 120	P
Cadmium	ug/L	1000	965		96	80 - 120	P
Chromium	ug/L	2000	1970		98	80 - 120	P
Lead	ug/L	5000	4900		98	80 - 120	P
Selenium	ug/L	10000	10800		108	80 - 120	P
Silver	ug/L	380	405		107	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2743

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169140BS Mercury	ug/L	4.0	4.08		102	80 - 120	CV