

Continuing Calibration Report for SequenceID : FC081225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069634.D

Aliphatic C9-C12	8084744.000	60.000	3.296	6.594	134745.733	144679.273	6.866
Aliphatic C12-C16	5438811.000	40.000	6.595	9.998	135970.275	143435.206	5.204
Aliphatic C16-C21	7579864.000	60.000	9.999	13.367	126331.067	133561.231	5.413
Aliphatic C21-C28	8897980.000	80.000	13.368	17.032	111224.750	118307.822	5.987
Aliphatic C28-C40	10856898.000	120.000	17.033	22.008	90474.150	95315.213	5.079
Aliphatic EPH	40858297.000	360.000	3.296	22.008	113495.269	120373.027	5.714