

Continuing Calibration Report for SequenceID : FD081225AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049615.D

Aromatic C10-C12	4939839.000	40.000	4.406	6.150	123495.975	107939.663	-14.412
Aromatic C12-C16	8456670.000	60.000	6.151	8.778	140944.500	123580.152	-14.051
Aromatic C16-C21	11415221.000	80.000	8.779	13.070	142690.263	126486.361	-12.811
Aromatic C21-C36	21792549.000	180.000	13.071	18.490	121069.717	112618.254	-7.505
Aromatic EPH	46604279.000	360.000	4.406	18.490	129456.331	117007.195	-10.640