

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC080125AL

Review By	yogesh	Review On	8/1/2025 10:37:24 AM
Supervise By	mohammad	Supervise On	8/5/2025 1:59:30 AM
SubDirectory	FC080125AL	HP Acquire Method	HP Processing Method FC080125AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC	PP24771		
Internal Standard/PEM			
ICV/I.BLK	PP24769,PP24774		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FC069556.D	01 Aug 2025 09:39		YP/AJ	Ok
2	I.BLK	I.BLK	FC069557.D	01 Aug 2025 10:20		YP/AJ	Ok
3	100 PPM ALIPHATIC HC	100 PPM ALIPHATIC HC	FC069558.D	01 Aug 2025 11:01		YP/AJ	Ok
4	50 PPM ALIPHATIC HC	50 PPM ALIPHATIC HC	FC069559.D	01 Aug 2025 11:43		YP/AJ	Ok
5	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069560.D	01 Aug 2025 12:24		YP/AJ	Ok
6	10 PPM ALIPHATIC HC	10 PPM ALIPHATIC HC	FC069561.D	01 Aug 2025 13:06		YP/AJ	Ok
7	5 PPM ALIPHATIC HC	5 PPM ALIPHATIC HC	FC069562.D	01 Aug 2025 13:47		YP/AJ	Ok
8	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069563.D	01 Aug 2025 14:28		YP/AJ	Ok
9	I.BLK	I.BLK	FC069564.D	01 Aug 2025 15:10		YP/AJ	Ok
10	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069565.D	01 Aug 2025 15:52		YP/AJ	Ok
11	Q2740-01	OR-02-07312025	FC069566.D	01 Aug 2025 16:37	Need 2X	YP/AJ	Dilution
12	Q2740-01D	Q2740-01D	FC069567.D	01 Aug 2025 17:19		YP/AJ	Ok
13	Q2740-01MS	OR-02-07312025MS	FC069568.D	01 Aug 2025 18:02	FC069566.D	YP/AJ	Ok,M
14	Q2740-01MSD	OR-02-07312025MSD	FC069569.D	01 Aug 2025 18:44	FC069566.D!FC069568.D	YP/AJ	Ok,M
15	I.BLK	I.BLK	FC069570.D	01 Aug 2025 19:26		YP/AJ	Ok
16	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069571.D	01 Aug 2025 20:09		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC081225AL

Review By	yogesh	Review On	8/12/2025 10:15:09 AM
Supervise By	mohammad	Supervise On	8/13/2025 9:40:39 AM
SubDirectory	FC081225AL	HP Acquire Method	HP Processing Method FC080125AL

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC Internal Standard/PEM	PP24771
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FC069632.D	12 Aug 2025 09:33		YP/AJ	Ok
2	I.BLK	I.BLK	FC069633.D	12 Aug 2025 10:17		YP/AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069634.D	12 Aug 2025 11:03		YP/AJ	Ok
4	PB169209BL	PB169209BL	FC069635.D	12 Aug 2025 11:48		YP/AJ	Ok
5	PB169209BS	PB169209BS	FC069636.D	12 Aug 2025 12:34	naphthalene break down - 0.356, 2-methylnepthalene break down - 0.156	YP/AJ	Ok
6	PB169209BSD	PB169209BSD	FC069637.D	12 Aug 2025 13:21	naphthalene break down - 0.399, 2-methylnepthalene break down - 0.209	YP/AJ	Ok
7	Q2743-01	WC1	FC069638.D	12 Aug 2025 14:10		YP/AJ	Ok
8	Q2743-01D	Q2743-01D	FC069639.D	12 Aug 2025 15:01		YP/AJ	Ok
9	Q2743-01MS	WC1MS	FC069640.D	12 Aug 2025 15:46	FC069638.D	YP/AJ	Ok
10	Q2743-01MSD	WC1MSD	FC069641.D	12 Aug 2025 16:32	FC069638.D!FC069640.D	YP/AJ	Ok
11	I.BLK	I.BLK	FC069642.D	12 Aug 2025 18:04		YP/AJ	Ok
12	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069643.D	12 Aug 2025 18:50		YP/AJ	Ok
13	PB169207BL	PB169207BL	FC069644.D	12 Aug 2025 20:22		YP/AJ	Ok
14	PB169207BS	PB169207BS	FC069645.D	12 Aug 2025 21:08		YP/AJ	Ok
15	PB169207BSD	PB169207BSD	FC069646.D	12 Aug 2025 21:53		YP/AJ	Ok
16	Q2830-03	BIN0009-DRIVEWAY-T	FC069647.D	12 Aug 2025 22:41		YP/AJ	Ok

Instrument ID: FID_C

Daily Analysis Runlog For Sequence/QC Batch ID # FC081225AL

Review By	yogesh	Review On	8/12/2025 10:15:09 AM
Supervise By	mohammad	Supervise On	8/13/2025 9:40:39 AM
SubDirectory	FC081225AL	HP Acquire Method	HP Processing Method FC080125AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC Internal Standard/PEM	PP24771		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774		

17	Q2830-03D	Q2830-03D	FC069648.D	12 Aug 2025 23:27		YP/AJ	Ok
18	Q2830-03MS	BIN0009-DRIVEWAY-T	FC069649.D	13 Aug 2025 00:13	FC069647.D	YP/AJ	Ok
19	Q2830-03MSD	BIN0009-DRIVEWAY-T	FC069650.D	13 Aug 2025 01:50	FC069647.D!FC069649.D	YP/AJ	Ok
20	Q2830-04	BIN0009-DRIVEWAY-T	FC069651.D	13 Aug 2025 02:34		YP/AJ	Ok
21	I.BLK	I.BLK	FC069652.D	13 Aug 2025 04:02		YP/AJ	Ok
22	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FC069653.D	13 Aug 2025 04:46		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_D

Daily Analysis Runlog For Sequence/QC Batch ID # FD081125AR

Review By	yogesh	Review On	8/11/2025 12:27:21 PM
Supervise By	mohammad	Supervise On	8/13/2025 1:17:39 AM
SubDirectory	FD081125AR	HP Acquire Method	HP Processing Method FD081125AR
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24500,PP24502,PP24503,PP24504,PP24505		
CCC	PP24503		
Internal Standard/PEM			
ICV/I.BLK	PP24501,PP24506		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FD049603.D	11 Aug 2025 08:46		YP/AJ	Ok
2	I.BLK	I.BLK	FD049604.D	11 Aug 2025 09:29		YP/AJ	Ok
3	100 PPM AROMATIC HC	100 PPM AROMATIC HC	FD049605.D	11 Aug 2025 10:39		YP/AJ	Ok
4	50 PPM AROMATIC HC	50 PPM AROMATIC HC	FD049606.D	11 Aug 2025 11:22		YP/AJ	Ok
5	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD049607.D	11 Aug 2025 12:06		YP/AJ	Ok
6	10 PPM AROMATIC HC	10 PPM AROMATIC HC	FD049608.D	11 Aug 2025 12:51		YP/AJ	Ok
7	5 PPM AROMATIC HC	5 PPM AROMATIC HC	FD049609.D	11 Aug 2025 13:35		YP/AJ	Ok
8	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD049610.D	11 Aug 2025 14:20		YP/AJ	Ok
9	I.BLK	I.BLK	FD049611.D	11 Aug 2025 15:05		YP/AJ	Ok
10	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD049612.D	11 Aug 2025 15:52		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_D

Daily Analysis Runlog For Sequence/QC Batch ID # FD081225AR

Review By	yogesh	Review On	8/12/2025 10:17:06 AM
Supervise By	mohammad	Supervise On	8/13/2025 9:40:46 AM
SubDirectory	FD081225AR	HP Acquire Method	HP Processing Method FD081125AR

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24500,PP24502,PP24503,PP24504,PP24505
CCC Internal Standard/PEM	PP24503
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24501,PP24506

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FD049613.D	12 Aug 2025 09:33		YP/AJ	Ok
2	I.BLK	I.BLK	FD049614.D	12 Aug 2025 10:17		YP/AJ	Ok
3	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD049615.D	12 Aug 2025 11:03		YP/AJ	Ok
4	PB169209BL	PB169209BL	FD049616.D	12 Aug 2025 11:48		YP/AJ	Ok
5	PB169209BS	PB169209BS	FD049617.D	12 Aug 2025 12:34	naphthalene break down - 0.356, 2-methylnepthalene break down - 0.156	YP/AJ	Ok,M
6	PB169209BSD	PB169209BSD	FD049618.D	12 Aug 2025 13:21	naphthalene break down - 0.399, 2-methylnepthalene break down - 0.209	YP/AJ	Ok,M
7	Q2743-01	WC1	FD049619.D	12 Aug 2025 14:10		YP/AJ	Ok
8	Q2743-01D	Q2743-01D	FD049620.D	12 Aug 2025 15:01		YP/AJ	Ok
9	Q2743-01MS	WC1MS	FD049621.D	12 Aug 2025 15:46	FD049619.D	YP/AJ	Ok
10	Q2743-01MSD	WC1MSD	FD049622.D	12 Aug 2025 16:32	FD049619.D!FD049621.D	YP/AJ	Ok,M
11	I.BLK	I.BLK	FD049623.D	12 Aug 2025 18:04		YP/AJ	Ok
12	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FD049624.D	12 Aug 2025 18:50		YP/AJ	Ok

M : Manual Integration