

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q2743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112669.D **Time Analyzed:** 17:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.754	4.655	4.855	514.840	500.000	3.0
Aroclor-1016-2	4.772	4.674	4.874	510.810	500.000	2.2
Aroclor-1016-3	4.829	4.731	4.931	501.300	500.000	0.3
Aroclor-1016-4	4.949	4.851	5.051	504.280	500.000	0.9
Aroclor-1016-5	5.206	5.108	5.308	529.170	500.000	5.8
Aroclor-1260-1	6.242	6.145	6.345	516.150	500.000	3.2
Aroclor-1260-2	6.432	6.333	6.533	510.260	500.000	2.1
Aroclor-1260-3	6.797	6.700	6.900	540.320	500.000	8.1
Aroclor-1260-4	7.057	6.959	7.159	512.110	500.000	2.4
Aroclor-1260-5	7.300	7.203	7.403	512.920	500.000	2.6
Decachlorobiphenyl	8.690	8.593	8.793	51.050	50.000	2.1
Tetrachloro-m-xylene	3.669	3.569	3.769	50.460	50.000	0.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q2743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112669.D **Time Analyzed:** 17:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.735	4.638	4.838	469.070	500.000	-6.2
Aroclor-1016-2	4.754	4.656	4.856	468.840	500.000	-6.2
Aroclor-1016-3	4.928	4.831	5.031	469.990	500.000	-6.0
Aroclor-1016-4	4.971	4.873	5.073	474.700	500.000	-5.1
Aroclor-1016-5	5.183	5.085	5.285	482.200	500.000	-3.6
Aroclor-1260-1	6.210	6.113	6.313	490.920	500.000	-1.8
Aroclor-1260-2	6.399	6.302	6.502	489.550	500.000	-2.1
Aroclor-1260-3	6.550	6.453	6.653	501.170	500.000	0.2
Aroclor-1260-4	7.019	6.923	7.123	501.040	500.000	0.2
Aroclor-1260-5	7.261	7.164	7.364	516.750	500.000	3.4
Decachlorobiphenyl	8.635	8.540	8.740	49.030	50.000	-1.9
Tetrachloro-m-xylene	3.661	3.562	3.762	47.760	50.000	-4.5