

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q2743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112684.D **Time Analyzed:** 23:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.753	4.655	4.855	515.160	500.000	3.0
Aroclor-1016-2	4.772	4.674	4.874	509.870	500.000	2.0
Aroclor-1016-3	4.829	4.731	4.931	504.170	500.000	0.8
Aroclor-1016-4	4.948	4.851	5.051	504.910	500.000	1.0
Aroclor-1016-5	5.204	5.108	5.308	519.150	500.000	3.8
Aroclor-1260-1	6.242	6.145	6.345	462.370	500.000	-7.5
Aroclor-1260-2	6.432	6.333	6.533	471.310	500.000	-5.7
Aroclor-1260-3	6.797	6.700	6.900	528.970	500.000	5.8
Aroclor-1260-4	7.056	6.959	7.159	476.180	500.000	-4.8
Aroclor-1260-5	7.299	7.203	7.403	457.960	500.000	-8.4
Decachlorobiphenyl	8.689	8.593	8.793	44.870	50.000	-10.3
Tetrachloro-m-xylene	3.669	3.569	3.769	48.780	50.000	-2.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q2743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/01/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112684.D **Time Analyzed:** 23:25

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.735	4.638	4.838	470.850	500.000	-5.8
Aroclor-1016-2	4.753	4.656	4.856	472.360	500.000	-5.5
Aroclor-1016-3	4.928	4.831	5.031	478.280	500.000	-4.3
Aroclor-1016-4	4.971	4.873	5.073	476.770	500.000	-4.6
Aroclor-1016-5	5.182	5.085	5.285	500.310	500.000	0.1
Aroclor-1260-1	6.210	6.113	6.313	476.650	500.000	-4.7
Aroclor-1260-2	6.398	6.302	6.502	474.650	500.000	-5.1
Aroclor-1260-3	6.549	6.453	6.653	481.070	500.000	-3.8
Aroclor-1260-4	7.019	6.923	7.123	470.950	500.000	-5.8
Aroclor-1260-5	7.261	7.164	7.364	475.840	500.000	-4.8
Decachlorobiphenyl	8.635	8.540	8.740	47.230	50.000	-5.5
Tetrachloro-m-xylene	3.661	3.562	3.762	47.890	50.000	-4.2