

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080425\
 Data File : BF143290.D
 Acq On : 04 Aug 2025 12:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 12:25:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2	1,4-Dioxane	0.598	0.635	-6.2	74	-0.03
3	Pyridine	1.552	1.616	-4.1	73	-0.02
4	n-Nitrosodimethylamine	0.801	0.828	-3.4	71	-0.02
5 S	2-Fluorophenol	1.264	1.304	-3.2	74	0.00
6	Aniline	2.217	2.270	-2.4	72	0.00
7 S	Phenol-d6	1.591	1.645	-3.4	74	0.00
8	2-Chlorophenol	1.325	1.362	-2.8	72	0.00
9	Benzaldehyde	1.193	1.312	-10.0	63	0.00
10 C	Phenol	1.733	1.905	-9.9	77	0.00
11	bis(2-Chloroethyl)ether	1.327	1.390	-4.7	74	0.00
12	1,3-Dichlorobenzene	1.439	1.494	-3.8	74	0.00
13 C	1,4-Dichlorobenzene	1.449	1.494	-3.1	73	0.00
14	1,2-Dichlorobenzene	1.378	1.419	-3.0	74	0.00
15	Benzyl Alcohol	1.215	1.223	-0.7	70	0.00
16	2,2'-oxybis(1-Chloropropane	2.461	2.560	-4.0	74	0.00
17	2-Methylphenol	1.114	1.137	-2.1	72	0.00
18	Hexachloroethane	0.502	0.529	-5.4	74	0.00
19 P	n-Nitroso-di-n-propylamine	1.021	1.014	0.7	72	0.00
20	3+4-Methylphenols	1.351	1.378	-2.0	72	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	70	0.00
22	Acetophenone	0.474	0.474	0.0	73	0.00
23 S	Nitrobenzene-d5	0.401	0.415	-3.5	73	0.00
24	Nitrobenzene	0.374	0.391	-4.5	73	0.00
25	Isophorone	0.708	0.720	-1.7	73	0.00
26 C	2-Nitrophenol	0.162	0.180	-11.1	73	0.00
27	2,4-Dimethylphenol	0.331	0.332	-0.3	71	0.00
28	bis(2-Chloroethoxy)methane	0.425	0.436	-2.6	74	0.00
29 C	2,4-Dichlorophenol	0.279	0.287	-2.9	73	0.00
30	1,2,4-Trichlorobenzene	0.302	0.306	-1.3	73	0.00
31	Naphthalene	0.985	1.008	-2.3	74	0.00
32	Benzoic acid	0.169	0.161	4.7	63	0.00
33	4-Chloroaniline	0.402	0.413	-2.7	74	0.00
34 C	Hexachlorobutadiene	0.184	0.186	-1.1	72	0.00
35	Caprolactam	0.084	0.088	-4.8	70	-0.01
36 C	4-Chloro-3-methylphenol	0.303	0.305	-0.7	71	0.00
37	2-Methylnaphthalene	0.599	0.610	-1.8	73	0.00
38	1-Methylnaphthalene	0.617	0.629	-1.9	73	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
40	1,2,4,5-Tetrachlorobenzene	0.577	0.610	-5.7	74	0.00
41 P	Hexachlorocyclopentadiene	0.376	0.356	5.3	64	0.00
42 S	2,4,6-Tribromophenol	0.207	0.215	-3.9	69	0.00
43 C	2,4,6-Trichlorophenol	0.400	0.395	1.3	69	0.00
44	2,4,5-Trichlorophenol	0.400	0.412	-3.0	68	0.00
45 S	2-Fluorobiphenyl	1.462	1.487	-1.7	74	0.00
46	1,1'-Biphenyl	1.560	1.631	-4.6	74	0.00
47	2-Chloronaphthalene	1.155	1.190	-3.0	72	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080425\
 Data File : BF143290.D
 Acq On : 04 Aug 2025 12:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 04 12:25:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 17 15:14:05 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.362	0.399	-10.2	71	0.00
49	Acenaphthylene	1.935	2.001	-3.4	72	0.00
50	Dimethylphthalate	1.304	1.359	-4.2	71	0.00
51	2,6-Dinitrotoluene	0.265	0.284	-7.2	69	0.00
52 C	Acenaphthene	1.144	1.188	-3.8	73	0.00
53	3-Nitroaniline	0.310	0.324	-4.5	69	0.00
54 P	2,4-Dinitrophenol	0.098	0.104	-6.1	67	0.00
55	Dibenzofuran	1.698	1.735	-2.2	71	0.00
56 P	4-Nitrophenol	0.232	0.275	-18.5	75	0.01
57	2,4-Dinitrotoluene	0.333	0.363	-9.0	69	0.00
58	Fluorene	1.274	1.320	-3.6	73	0.00
59	2,3,4,6-Tetrachlorophenol	0.331	0.342	-3.3	69	0.00
60	Diethylphthalate	1.289	1.338	-3.8	71	0.00
61	4-Chlorophenyl-phenylether	0.635	0.664	-4.6	73	0.00
62	4-Nitroaniline	0.266	0.285	-7.1	68	0.00
63	Azobenzene	1.343	1.419	-5.7	72	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
65	4,6-Dinitro-2-methylphenol	0.094	0.101	-7.4	69	0.00
66 c	n-Nitrosodiphenylamine	0.704	0.722	-2.6	71	0.00
67	4-Bromophenyl-phenylether	0.238	0.247	-3.8	71	0.00
68	Hexachlorobenzene	0.249	0.251	-0.8	69	0.00
69	Atrazine	0.194	0.211	-8.8	71	0.00
70 C	Pentachlorophenol	0.147	0.134	8.8	60	0.00
71	Phenanthrene	1.062	1.085	-2.2	70	0.00
72	Anthracene	1.083	1.123	-3.7	71	0.00
73	Carbazole	0.946	0.979	-3.5	69	0.00
74	Di-n-butylphthalate	1.070	1.194	-11.6	73	0.00
75 C	Fluoranthene	0.975	1.058	-8.5	73	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
77	Benzidine	0.771	0.542	29.7#	43#	0.00
78	Pyrene	1.707	1.713	-0.4	75	0.00
79 S	Terphenyl-d14	1.344	1.333	0.8	76	0.00
80	Butylbenzylphthalate	0.523	0.618	-18.2	85	0.00
81	Benzo(a)anthracene	1.339	1.356	-1.3	79	0.00
82	3,3'-Dichlorobenzidine	0.446	0.462	-3.6	81	0.00
83	Chrysene	1.204	1.248	-3.7	77	0.00
84	Bis(2-ethylhexyl)phthalate	0.791	0.879	-11.1	83	0.00
85 c	Di-n-octyl phthalate	1.434	1.428	0.4	77	0.00
86 I	Perylene-d12	1.000	1.000	0.0	76	0.01
87	Indeno(1,2,3-cd)pyrene	1.410	1.475	-4.6	79	0.01
88	Benzo(b)fluoranthene	1.222	1.303	-6.6	86	0.00
89	Benzo(k)fluoranthene	1.090	1.036	5.0	69	0.00
90 C	Benzo(a)pyrene	1.126	1.162	-3.2	79	0.00
91	Dibenzo(a,h)anthracene	1.148	1.210	-5.4	80	0.01
92	Benzo(g,h,i)perylene	1.110	1.206	-8.6	82	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080425\
Data File : BF143290.D
Acq On : 04 Aug 2025 12:04
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Aug 04 12:25:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF071725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 17 15:14:05 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0