

Surrogate Summary

SW-846

SDG No.: Q2743

Client: G Environmental

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169095BL	PB169095BL	2-Fluorophenol	150	123	82		30 (18)	130 (112)
		Phenol-d6	150	124	83		30 (15)	130 (107)
		Nitrobenzene-d5	100	85.1	85		30 (18)	130 (107)
		2-Fluorobiphenyl	100	82.1	82		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	127	84		30 (10)	130 (116)
		Terphenyl-d14	100	82.4	82		30 (10)	130 (105)
PB169095BS	PB169095BS	2-Fluorophenol	150	119	79		30 (18)	130 (112)
		Phenol-d6	150	119	80		30 (15)	130 (107)
		Nitrobenzene-d5	100	82.0	82		30 (18)	130 (107)
		2-Fluorobiphenyl	100	79.4	79		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	130	86		30 (10)	130 (116)
		Terphenyl-d14	100	80.3	80		30 (10)	130 (105)
Q2734-01MS	VNJ-253MS	2-Fluorophenol	150	80.6	54		30 (18)	130 (112)
		Phenol-d6	150	83.6	56		30 (15)	130 (107)
		Nitrobenzene-d5	100	56.4	56		30 (18)	130 (107)
		2-Fluorobiphenyl	100	59.0	59		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	71.9	48		30 (10)	130 (116)
		Terphenyl-d14	100	47.0	47		30 (10)	130 (105)
Q2734-01MSD	VNJ-253MSD	2-Fluorophenol	150	87.2	58		30 (18)	130 (112)
		Phenol-d6	150	89.9	60		30 (15)	130 (107)
		Nitrobenzene-d5	100	60.6	61		30 (18)	130 (107)
		2-Fluorobiphenyl	100	64.5	64		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	76.0	51		30 (10)	130 (116)
		Terphenyl-d14	100	50.7	51		30 (10)	130 (105)
Q2743-01	WC1	2-Fluorophenol	150	76.2	51		30 (18)	130 (112)
		Phenol-d6	150	78.4	52		30 (15)	130 (107)
		Nitrobenzene-d5	100	54.2	54		30 (18)	130 (107)
		2-Fluorobiphenyl	100	54.7	55		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	74.2	49		30 (10)	130 (116)
		Terphenyl-d14	100	40.5	41		30 (10)	130 (105)