

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2743

Client ID : SSTDCCC040

Date Analyzed: 08/04/2025

Lab File ID: BF143290.D

Time Analyzed: 12:04

Instrument ID: BNA\_F

GC Column: DB-UI ID: 0.18 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #  | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|-------|---------------------|--------|
| 12 HOUR STD    | 116257              | 6.963 | 448345              | 8.25  | 222822              | 10.00  |
| UPPER LIMIT    | 232514              | 7.463 | 896690              | 8.745 | 445644              | 10.498 |
| LOWER LIMIT    | 58128.5             | 6.463 | 224173              | 7.745 | 111411              | 9.498  |
| EPA SAMPLE NO. |                     |       |                     |       |                     |        |
| 01 WC1         | 122191              | 6.96  | 459211              | 8.24  | 223785              | 9.99   |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.