

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW072225

Review By	Semsettin Yesilyurt	Review On	7/24/2025 3:08:17 PM
Supervise By	Mahesh Dadoda	Supervise On	7/24/2025 3:08:52 PM
SubDirectory	VW072225	HP Acquire Method	MSVOA_W HP Processing Method 82w072225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134865 VP134866,VP134868,VP134869,VP134870,VP134873,VP134874
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133934 VP134875

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031890.D	22 Jul 2025 08:14		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VW031891.D	22 Jul 2025 09:05	Comp. #20 is on Linear Regression	SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VW031892.D	22 Jul 2025 09:37	Comp. #13 is on Quadratic Regression	SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VW031893.D	22 Jul 2025 10:17		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VW031894.D	22 Jul 2025 10:39		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VW031895.D	22 Jul 2025 11:18		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VW031896.D	22 Jul 2025 12:00		SY/MD	Ok,M
8	VIBLK	VIBLK	VW031897.D	22 Jul 2025 12:21		SY/MD	Ok
9	VSTDICV050	ICVVW072225	VW031898.D	22 Jul 2025 14:47		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_W

Daily Analysis Runlog For Sequence/QC Batch ID # VW080125

Review By	Maresh Dadoda	Review On	8/4/2025 1:49:28 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/4/2025 2:29:21 PM
SubDirectory	VW080125	HP Acquire Method	MSVOA_W HP Processing Method 82w072225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134960
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134962,VP134963 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VW031970.D	01 Aug 2025 07:52		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VW031971.D	01 Aug 2025 08:21		SY/MD	Ok,M
3	VW0801SBL01	VW0801SBL01	VW031972.D	01 Aug 2025 09:21		SY/MD	Ok
4	VW0801SBS01	VW0801SBS01	VW031973.D	01 Aug 2025 09:49		SY/MD	Ok,M
5	VW0801SBSD01	VW0801SBSD01	VW031974.D	01 Aug 2025 10:12		SY/MD	Ok,M
6	Q2741-01	TR-05-07312025	VW031975.D	01 Aug 2025 11:12	vial-A Internal standard fail;Surrogate fail	SY/MD	ReRun
7	Q2740-01	OR-02-07312025	VW031976.D	01 Aug 2025 11:34	vial-A	SY/MD	Ok
8	Q2743-01	WC1	VW031977.D	01 Aug 2025 11:56	vial-A	SY/MD	Ok
9	Q2747-01	OU4-TS-GRILLO-OG2	VW031978.D	01 Aug 2025 12:18	vial-A Surrogate fail	SY/MD	ReRun
10	Q2747-03	OU4-TS-GRILLO-OG3	VW031979.D	01 Aug 2025 12:40	vial-A Surrogate fail	SY/MD	ReRun
11	Q2747-05	OU4-TS-GRILLO-OG4	VW031980.D	01 Aug 2025 13:02	vial-A Surrogate fail	SY/MD	ReRun
12	Q2747-07	OU4-TS-GRILLO-TSCF	VW031981.D	01 Aug 2025 13:24	vial-A Surrogate fail	SY/MD	ReRun
13	Q2747-09	OU4-TS-GRILLO-TSCF	VW031982.D	01 Aug 2025 13:47	vial-A Surrogate fail	SY/MD	ReRun
14	Q2741-01	TR-05-07312025	VW031983.D	01 Aug 2025 14:09	vial-B	SY/MD	Ok
15	VIBLK	VIBLK	VW031984.D	01 Aug 2025 14:35		SY/MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VW031985.D	01 Aug 2025 14:57		SY/MD	Ok,M

M : Manual Integration