

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2743
 Lab File ID: VW031971.D Date Analyzed: 08/01/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:21
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	217356	7.96	389384	8.85	352859	11.63
UPPER LIMIT	434712	8.459	778768	9.349	705718	12.129
LOWER LIMIT	108678	7.459	194692	8.349	176430	11.129
EPA SAMPLE NO.						
WC1	174433	7.96	365996	8.86	361399	11.64
VW0801SBL01	165069	7.96	350388	8.85	344045	11.64
VW0801SBS01	204409	7.96	389040	8.85	355301	11.63
VW0801SBSD01	220615	7.96	398144	8.86	370224	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.