

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02  
 Lab Code: ACE SDG No.: Q2744  
 Instrument ID: MSVOA\_X Calibration Date/Time: 08/01/2025 10:22  
 Lab File ID: VX047202.D Init. Calib. Date(s): 07/21/2025 07/21/2025  
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 11:32  
 GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND           | Amount  | Conc.  | %Dev     | MAX%D |
|--------------------|---------|--------|----------|-------|
| Text butyl alcohol | 250.000 | 235.06 | 5.978003 | 20    |