



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX080125\
Data File : VX047201.D
Acq On : 01 Aug 2025 09:10
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Title : SW846 8260
Last Update : Tue Jul 22 03:59:51 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.2	27525	PASS
75	95	30	60	51.7	74141	PASS
95	95	100	100	100.0	143379	PASS
96	95	5	9	6.3	9087	PASS
173	174	0.00	2	0.9	958	PASS
174	95	50	100	73.3	105072	PASS
175	174	5	9	7.3	7670	PASS
176	174	95	101	99.0	104043	PASS
177	176	5	9	6.5	6777	PASS