

Surrogate Summary

SDG No.: Q2744

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2744-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	49.3	99		74	125
		Dibromofluoromethane	50	50.3	101		75	124
		Toluene-d8	50	46.7	93		86	113
		4-Bromofluorobenzene	50	50.8	102		77	121
Q2744-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	59.4	119		74	125
		Dibromofluoromethane	50	58.9	118		75	124
		Toluene-d8	50	54.8	110		86	113
		4-Bromofluorobenzene	50	60.1	120		77	121
Q2744-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	53.8	108		74	125
		Dibromofluoromethane	50	48.9	98		75	124
		Toluene-d8	50	54.9	110		86	113
		4-Bromofluorobenzene	50	52.7	105		77	121
Q2744-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	53.7	107		74	125
		Dibromofluoromethane	50	54.0	108		75	124
		Toluene-d8	50	49.6	99		86	113
		4-Bromofluorobenzene	50	53.9	108		77	121
VX0801WBL01	VX0801WBL01	1,2-Dichloroethane-d4	50	50.2	100		74	125
		Dibromofluoromethane	50	50.5	101		75	124
		Toluene-d8	50	47.2	94		86	113
		4-Bromofluorobenzene	50	51.0	102		77	121
VX0801WBS01	VX0801WBS01	1,2-Dichloroethane-d4	50	53.2	106		74	125
		Dibromofluoromethane	50	54.7	109		75	124
		Toluene-d8	50	50.4	101		86	113
		4-Bromofluorobenzene	50	54.9	110		77	121
VX0804WBL01	VX0804WBL01	1,2-Dichloroethane-d4	50	54.9	110		74	125
		Dibromofluoromethane	50	49.7	99		75	124
		Toluene-d8	50	54.5	109		86	113
		4-Bromofluorobenzene	50	53.4	107		77	121
VX0804WBS01	VX0804WBS01	1,2-Dichloroethane-d4	50	56.6	113		74	125
		Dibromofluoromethane	50	59.9	120		75	124
		Toluene-d8	50	55.3	111		86	113
		4-Bromofluorobenzene	50	58.8	118		77	121
VX0804WBSD01	VX0804WBSD01	1,2-Dichloroethane-d4	50	48.7	97		74	125
		Dibromofluoromethane	50	49.2	98		75	124
		Toluene-d8	50	45.9	92		86	113
		4-Bromofluorobenzene	50	48.6	97		77	121