

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q2744
 Lab File ID: VX047202.D Date Analyzed: 08/01/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	249274	5.56	439479	6.76	396034	10.05
UPPER LIMIT	498548	6.056	878958	7.263	792068	10.549
LOWER LIMIT	124637	5.056	219740	6.263	198017	9.549
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	283674	5.56	488674	6.77	440161	10.06
STORAGE-BLANK-WATER-REF-5	262356	5.57	449740	6.77	402679	10.06
STORAGE-BLANK-SAMPLE-MANAGEMENT	289476	5.57	495196	6.77	453951	10.06
VX0801WBL01	289784	5.56	493740	6.76	446060	10.05
VX0801WBS01	260557	5.56	443908	6.76	390421	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.