

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q2744
 Lab File ID: VX047202.D Date Analyzed: 08/01/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	195271	12.018				
UPPER LIMIT	390542	12.518				
LOWER LIMIT	97635.5	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	222884	12.02				
STORAGE-BLANK-WATER-REF-5	207141	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	225621	12.02				
VX0801WBL01	227682	12.02				
VX0801WBS01	195696	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.