

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3980	4000	100	90 - 110	P	08/04/2025	11:41	LB136697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	102	100	102	80 - 120	P	08/04/2025	11:45	LB136697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4860	5000	97	90 - 110	P	08/04/2025	12:25	LB136697
CCV02	Iron	4930	5000	99	90 - 110	P	08/04/2025	13:37	LB136697
CCV03	Iron	4990	5000	100	90 - 110	P	08/04/2025	14:43	LB136697
CCV04	Iron	4900	5000	98	90 - 110	P	08/04/2025	15:48	LB136697
CCV05	Iron	5040	5000	101	90 - 110	P	08/04/2025	16:51	LB136697
CCV06	Iron	5210	5000	104	90 - 110	P	08/04/2025	17:41	LB136697
CCV07	Iron	5150	5000	103	90 - 110	P	08/04/2025	17:58	LB136697

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4000	4000	100	90 - 110	P	08/06/2025	11:55	LB136729

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	108	100	108	80 - 120	P	08/06/2025	12:09	LB136729

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4990	5000	100	90 - 110	P	08/06/2025	12:47	LB136729
CCV02	Iron	4940	5000	99	90 - 110	P	08/06/2025	13:37	LB136729
CCV03	Iron	5270	5000	106	90 - 110	P	08/06/2025	14:49	LB136729
