



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	111	100	111	65 - 135	P	08/04/2025	11:57	LB136697
CRI01	Iron	97.9	100	98	65 - 135	P	08/06/2025	12:26	LB136729