

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc.

SDG No.: Q2745

Contract: TETR06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	98400	101000	97	85600	116500	08/04/2025	12:01	LB136697
ICSAB01	Iron	101000	99300	102	84400	114500	08/04/2025	12:05	LB136697
ICSA01	Iron	102000	101000	101	85600	116500	08/06/2025	12:31	LB136729
ICSAB01	Iron	101000	99300	102	84400	114500	08/06/2025	12:35	LB136729