

metals  
- 5a -  
**MATRIX SPIKE SUMMARY**

**client:** Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q2745  
**contract:** TETR06 **lab code:** ACE  
**matrix:** Water **sample id:** Q2756-01 **client id:** 71725MS  
**Percent Solids for Sample:** NA **Spiked ID:** Q2756-01MS **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit %R | Spiked<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|------------------|---|------------------|---|----------------|---------------|------|---|
| Iron    | ug/L  | 87 - 115               | 6750             |   | 5880             |   | 1500           | 58            | N    | P |

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

**client:** Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q2745  
**contract:** TETR06 **lab code:** ACE  
**matrix:** Water **sample id:** Q2756-01 **client id:** 71725MSD  
**Percent Solids for Sample:** NA **Spiked ID:** Q2756-01MSD **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit %R | MSD<br>Result | C | Sample<br>Result | C | Spike<br>Added | %<br>Recovery | Qual | M |
|---------|-------|------------------------|---------------|---|------------------|---|----------------|---------------|------|---|
| Iron    | ug/L  | 87 - 115               | 7070          |   | 5880             |   | 1500           | 79            | N    | P |