

## Surrogate Summary

SW-846

SDG No.: Q2745

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

| Lab Sample ID | Client ID          | Parameter               | Spike (PPM) | Result (PPM) | Recovery (%) | Qual | Limits (%) |      |
|---------------|--------------------|-------------------------|-------------|--------------|--------------|------|------------|------|
|               |                    |                         |             |              |              |      | Low        | High |
| PB169094BL    | PB169094BL         | 2-Methylnaphthalene-d10 | 0.4         | 0.31         | 78           |      | 30         | 150  |
|               |                    | Fluoranthene-d10        | 0.4         | 0.32         | 81           |      | 30         | 150  |
|               |                    | Nitrobenzene-d5         | 0.4         | 0.35         | 88           |      | 55         | 111  |
|               |                    | 2-Fluorobiphenyl        | 0.4         | 0.38         | 94           |      | 53         | 106  |
|               |                    | Terphenyl-d14           | 0.4         | 0.39         | 97           |      | 58         | 132  |
| PB169094BS    | PB169094BS         | 2-Methylnaphthalene-d10 | 0.4         | 0.34         | 85           |      | 30         | 150  |
|               |                    | Fluoranthene-d10        | 0.4         | 0.31         | 77           |      | 30         | 150  |
|               |                    | Nitrobenzene-d5         | 0.4         | 0.35         | 87           |      | 55         | 111  |
|               |                    | 2-Fluorobiphenyl        | 0.4         | 0.41         | 102          |      | 53         | 106  |
|               |                    | Terphenyl-d14           | 0.4         | 0.36         | 90           |      | 58         | 132  |
| PB169094BSD   | PB169094BSD        | 2-Methylnaphthalene-d10 | 0.4         | 0.34         | 84           |      | 30         | 150  |
|               |                    | Fluoranthene-d10        | 0.4         | 0.30         | 75           |      | 30         | 150  |
|               |                    | Nitrobenzene-d5         | 0.4         | 0.35         | 88           |      | 55         | 111  |
|               |                    | 2-Fluorobiphenyl        | 0.4         | 0.40         | 100          |      | 53         | 106  |
|               |                    | Terphenyl-d14           | 0.4         | 0.37         | 92           |      | 58         | 132  |
| Q2745-01      | RW8-SP100-20250731 | 2-Methylnaphthalene-d10 | 0.4         | 0.32         | 79           |      | 30         | 150  |
|               |                    | Fluoranthene-d10        | 0.4         | 0.38         | 94           |      | 30         | 150  |
|               |                    | Nitrobenzene-d5         | 0.4         | 0.34         | 85           |      | 55         | 111  |
|               |                    | 2-Fluorobiphenyl        | 0.4         | 0.38         | 96           |      | 53         | 106  |
|               |                    | Terphenyl-d14           | 0.4         | 0.46         | 114          |      | 58         | 132  |
| Q2745-02      | RW8-SP303-20250731 | 2-Methylnaphthalene-d10 | 0.4         | 0.30         | 75           |      | 30         | 150  |
|               |                    | Fluoranthene-d10        | 0.4         | 0.35         | 88           |      | 30         | 150  |
|               |                    | Nitrobenzene-d5         | 0.4         | 0.32         | 80           |      | 55         | 111  |
|               |                    | 2-Fluorobiphenyl        | 0.4         | 0.36         | 90           |      | 53         | 106  |
|               |                    | Terphenyl-d14           | 0.4         | 0.47         | 116          |      | 58         | 132  |