

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q2746
 Instrument ID: BNA_N Calibration Date/Time: 08/05/2025 09:35
 Lab File ID: BN037559.D Init. Calib. Date(s): 07/15/2025 07/15/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:36 16:14
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.515		-10.3	20.0
Fluoranthene-d10	1.060	0.889		-16.1	20.0
2-Fluorophenol	0.989	0.880		-11.0	20.0
Phenol-d6	1.241	1.098		-11.5	20.0
Nitrobenzene-d5	0.299	0.279		-6.7	20.0
2-Fluorobiphenyl	2.080	2.181		4.9	20.0
2,4,6-Tribromophenol	0.197	0.140		-28.9	20.0
Terphenyl-d14	0.859	0.738		-14.1	20.0
1,4-Dioxane	0.385	0.404		4.9	20.0

All other compounds must meet a minimum RRF of 0.010.