

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2746</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>08/05/2025 15:03</u>
Lab File ID:	<u>BN037568.D</u>	Init. Calib. Date(s):	<u>07/15/2025 07/15/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>12:36 16:14</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.574	0.505		-12.0	50.0
Fluoranthene-d10	1.060	0.905		-14.6	50.0
2-Fluorophenol	0.989	0.840		-15.1	50.0
Phenol-d6	1.241	1.017		-18.0	50.0
Nitrobenzene-d5	0.299	0.288		-3.7	50.0
2-Fluorobiphenyl	2.080	2.112		1.5	50.0
2,4,6-Tribromophenol	0.197	0.145		-26.4	50.0
Terphenyl-d14	0.859	0.819		-4.7	50.0
1,4-Dioxane	0.385	0.388		0.8	50.0

All other compounds must meet a minimum RRF of 0.010.