

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037497.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2746
DFTPP Injection Date: 07/15/2025
DFTPP Injection Time: 10:57

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	83.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037499.D	07/15/2025	12:36
SSTDICC0.2	SSTDICC0.2	BN037500.D	07/15/2025	13:12
SSTDICCC0.4	SSTDICCC0.4	BN037501.D	07/15/2025	13:49
SSTDICC0.8	SSTDICC0.8	BN037502.D	07/15/2025	14:25
SSTDICC1.6	SSTDICC1.6	BN037503.D	07/15/2025	15:01
SSTDICC3.2	SSTDICC3.2	BN037504.D	07/15/2025	15:38
SSTDICC5.0	SSTDICC5.0	BN037505.D	07/15/2025	16:14