

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037558.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q2746
DFTPP Injection Date: 08/05/2025
DFTPP Injection Time: 08:56

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.4 (0.9) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.3 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	90.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.5 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037559.D	08/05/2025	09:35
PB169094BL	PB169094BL	BN037560.D	08/05/2025	10:11
RW7-SP100-20250731	Q2746-01	BN037563.D	08/05/2025	11:59
RW7-SP201-20250731	Q2746-02	BN037564.D	08/05/2025	12:35
RW7-SP303-20250731	Q2746-03	BN037565.D	08/05/2025	13:11
PB169094BS	PB169094BS	BN037566.D	08/05/2025	13:47
PB169094BSD	PB169094BSD	BN037567.D	08/05/2025	14:23
SSTDCCC0.4EC	SSTDCCC0.4	BN037568.D	08/05/2025	15:03