



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Nobis Group

Project Name: Raymark Superfund Site

Project # N/A

Order ID # Q2747

Test Name: SPLP Mercury,SPLP MetalGroup3

A. Number of Samples and Date of Receipt:

5 Solid samples were received on 08/01/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide, Herbicide Group1, PCB, Pesticide-TCL, SPLP Extraction, SPLP Mercury, SPLP MetalGroup3, SVOCMS Group3 and VOCMS Group3. This data package contains results for SPLP Mercury,SPLP MetalGroup3.

C. Analytical Techniques:

The analysis of SPLP MetalGroup3 was based on method 6020B, digestion based on method 3010 (water). The analysis of SPLP Mercury was based on method 7470A and digestion was based on method 7470 (water).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate (OU4-TS-GRILLO-TSCP02-073125MSD) analysis met criteria for all compounds except for Arsenic due to Chemical Interference during digestion process.

The Matrix Spike (OU4-TS-GRILLO-TSCP02-073125MS) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to Chemical Interference during digestion process.

The Matrix Spike Duplicate (OU4-TS-GRILLO-TSCP02-073125MSD) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to Chemical Interference during digestion process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The Post Digest Spike (OU4-TS-GRILLO-TSCP02-073125A) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.



Q2747 SPLP Metals Group3 all samples with its Qcs diluted 5X Straight due to SPLP fluid which cannot be injected as is without dilution to avoid damage to detector of instrument.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

In analytical sequence LB136712 The Result Concentration outside limit for Selenium of CCB02 and CCB04 but, no any sample associated under these CCBs.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____