

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2747

MATRIX: Water

METHOD: 6020B,7470A

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Serial Dilution Summary (if applicable) Submitted.			✓
4. Laboratory Control Sample Summary (if applicable) Submitted.			✓
5. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Matrix Spike (OU4-TS-GRILLO-TSCP02-073125MS) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to Chemical Interference during digestion process. The Matrix Spike Duplicate (OU4-TS-GRILLO-TSCP02-073125MSD) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to Chemical Interference during digestion process.			
7. Sample Duplicate Analysis Met QC Criteria		✓	
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Duplicate (OU4-TS-GRILLO-TSCP02-073125MSD) analysis met criteria for all compounds except for Arsenic due to Chemical Interference during digestion process.			
8. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
9. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The Post Digest Spike (OU4-TS-GRILLO-TSCP02-073125A) analysis met criteria for all compounds except for Antimony, Arsenic, Barium, Chromium, Silver, Thallium, Vanadium and Zinc due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.

Q2747 SPLP Metals Group3 all samples diluted 5X Straight due to SPLP fluid which cannot be injected as is without dilution to avoid damage to detector of instrument.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

In analytical sequence LB136712 The Result Concentration outside limit for Selenium of CCB02 and CCB04 but, no any sample associated under these CCBs.

QA REVIEW

Date