

Initial and Continuing Calibration Verification

Client: Nobis Group

SDG No.: Q2747

Project: Raymark Superfund Site

RunNo.: LB136702

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.094	0.099	95	90-110	08/05/2025
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	08/05/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	08/05/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	08/05/2025