

Initial and Continuing Calibration Blank Summary

Client: Nobis Group

SDG No.: Q2747

Project: Raymark Superfund Site

RunNo.: LB136702

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	08/05/2025
Sample ID: CCB1 Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	08/05/2025
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	08/05/2025
Sample ID: CCB3 Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	08/05/2025